



User's Manual

Human CD19(Cluster of Differentiation 19) ELISA Kit



DEIA-NS2307-100



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 **Address: 45-1 Ramsey Road, Shirley, NY 11967, USA**

 **Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)**  **Fax: 1-631-938-8221**

 **Email: info@creative-diagnostics.com**  **Web: www.creative-diagnostics.com**

PRODUCT INFORMATION


Intended Use

In vitro quantitative determination of CD19 concentrations in serum, plasma, cell culture supernatant and other biological samples.

Storage

2-8 °C for 6 months

Typical Standard Curve



Performance Characteristics

Perform the stability test for the sealed kit at 37°C and 2-8°C and get relevant data.

Elisa kit(n=5)	37°C for 1 month	2-8°C for 6 months
Average (%)	80	95-100

Precision

Intra-Assay: CV<8%

Inter-Assay: CV<10%

Detection Range

0.156-10ng/ml

Sensitivity

< 0.094ng/ml

Specificity

Specifically recognize CD19, no obvious cross reaction with other analogues

Linearity

Dilute the sample with a certain amount of CD19 at 1:2, 1:4 and 1:8 to get the recovery range.

Sample	1:2	1:4	1:8
Serum(n=5)	85-105%	86-100%	80-100%
EDTA Plasma(n=5)	81-100%	80-96%	82-98%
Heparin Plasma(n=5)	86-103%	83-98%	86-101%

Recovery

Add a certain amount of CD19 into the sample. Calculate the recovery by comparing the measured value with the expected amount of CD19 in the sample.

Matrix	Recovery Range (%)	Average (%)
Serum(n=5)	85-103	97
EDTA Plasma(n=5)	85-105	95
Heparin Plasma(n=5)	85-101	95