



## User's Manual

# Avian Influenza Virus H7 Antibodies ELISA Kit



DEIA-NS2312-2



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

Assay kit to test avian Influenza virus H7 antibodies.

### Principles of Testing

This kit applies the Competitive-ELISA as the method for the in vitro qualitative detection of concentration of antibodies to Avian Influenza Virus H7 (AIV-H7) in samples. The ELISA Microtiter plate provided in this kit has been pre-coated with the AIV-H7 antigen. In the experiment, add control serum, samples and Antibody Working Solution into plate. AIV-H7 antibodies in the samples will compete with the antibody in the working solution to bind with the antigen pre-coated on the Microplate. Then wash to remove unbound antibodies and other components, add the HRP Conjugate to specifically bind with the compound of antibody and antigen on the Microplate and the unbound HRP Conjugate will be removed by washing. Add TMB substrate to the wells, it will react with the enzyme and become blue, the shade of color is of negative correlation with AIV-H7 antibody levels in the samples. At last, end the reaction by adding Stop Solution to produce a yellow product. Measure the absorbance value of each well by using a Microplate Reader with 450 nm wavelength, then we can judge whether AIV-H7 antibody exist in the sample.

### Reagents And Materials Provided

1. ELISA Microtiter plate, 96 wells
2. Antibody Working Solution, 6 mL
3. HRP Conjugate, 11 mL
4. 20xConcentrated Wash Buffer, 40 mL
5. Substrate Reagent A, 6 mL
6. Substrate Reagent B, 6 mL
7. Stop Solution, 6 mL
8. Positive Control, 1 mL
9. Negative Control, 1 mL
10. Plate Sealer, 3 pieces
11. Sealed Bag, 1 piece
12. Manual, 1 copy

**Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.**

### Materials Required But Not Supplied

Microplate Reader with 450nm wavelength filter or dual-wavelength (450/630nm)

High-precision transferpettor, EP tubes and disposable pipette tips

37° C incubator or water bath

Deionized or distilled water

Absorbent paper

## Storage

Store at 2-8°C. Avoid freeze.

Please store the opened plate at 2-8°C, the shelf life of the opened kit is up to 1 month.

Expiry date: expiration date is on the packing box.

## Specimen Collection And Preparation

**Serum/plasma:** Use the conventional method to prepare serum/plasma, the serum/plasma must be clear, no hemolysis and no pollution. Samples can be conserved at 2~8°C in 1 weeks, and it should be stored at -20°C for a long term storage.

**Yolk:** Take 2 mL of fresh yolk and add 2 mL of normal saline, oscillate to mix fully. Centrifuge at 3000 rpm for 15 min, take the supernatant for detection.

## Reagent Preparation

**Wash Buffer:** The 20xConcentrated Wash Buffer should be adjusted to room temperature before used, then dilute it with deionized or distilled water at 1:19.

## Assay Procedure

Restore all reagents and samples to room temperature before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming.

1. Number: Take out the Microtiter plate, set 2 wells for negative/positive control respectively. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2~8°C. Standard and Samples need test in duplicate.
2. Add sample: Add 50 µL of positive/negative control to the wells respectively, add 10 µL of sample and 40 µL of Wash Buffer to each sample well.
3. Incubate: Then add 50 µL of Antibody Working Solution to each well. Gently tap the plate to ensure thorough mixing, incubate at 37°C for 30 min.
4. Wash: Remove the liquid in each well. Immediately add 300 µL of Wash Buffer to each well and wash. Repeat wash procedure for 5 times, 30 s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
5. HRP conjugate: Add 100 µL of HRP Conjugate into each well, and incubate at 37°C for 30 min.
6. Wash: Repeat Step 4 for washing.
7. Color Development: Add 50 µL of Substrate Reagent A and 50 µL of Substrate Reagent B into each well, gently tap the plate to ensure thorough mixing, incubate for 15 min at 37°C. Protect from light.
8. Stop reaction: Add 50 µL of Stop Solution into each well, gently mix.
9. OD Measurement: Measure the absorbance value (A-value) of each well by using a Microplate Reader with



450 nm wavelength (use 630 nm as reference wavelength).

## Interpretation Of Results

1.  $PI = (1 - \text{Sample OD} / \text{Average of negative control ODs}) \times 100$ .
2. Positive result :  $PI \geq 50$
3. Negative result :  $PI < 50$ .
4. Unimmunized animal: positive result indicates that it may be infected with AIV-H7.
5. Immunized poultry: The antibody levels at the time of the sample were monitored and recorded, and the distribution of antibody levels and the trend of immune status of the flock were analyzed based on the results.

**Note: The negative result of this test suggests that the concentration of antibody for the tested samples is not enough, it is recommended that this animal should get immunized with corresponding vaccine.**

## Reference Values

Normally, the OD of negative control  $\geq 1.0$  and the A-value of positive control  $\leq 50\%$   $\times$  A-value of negative control.

## Sensitivity

sensitivity > 98%, specificity > 98%

## Precautions

1. Wear gloves and work clothes during experiment, and the disinfection and isolation system should be strictly executed. All the waste should be handled as contaminant.
2. The Stop Solution is corrosive, it should be avoided to contact with skin and clothing. Wash immediately with plenty of water if contact it carelessly.
3. FOR RESEARCH USE ONLY. ELISA Microtiter plate should be covered by plate sealer. Avoid the kit to strong light.
4. Concentrated Wash Solution at low temperature condition is easy to crystallize, it should be adjusted to room temperature in order to dissolve completely before use.
5. Each well must be filled with liquid when washing in order to prevent residual free enzyme.
6. The tested sample should be kept fresh.
7. The results shall depend on the readings of the Microplate Reader.
8. Each reagent is optimized for use in the CD kit. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other kits with different lot numbers.
9. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.

## Limitations

1. This test is only used as the qualitative detection of AIV-H7 antibody in serum and plasma samples of Poultry. A rough estimate (high, general, low) of the antibody concentration can be calculated according to the PI values.
2. The detection results of this kit are only for clinical reference. For confirmation of the diagnosis, please combine the clinical symptoms and other methods of detection, this detection cannot be used as the only criteria for clinical diagnosis.
3. All high sensitivity immune experiment system exists potential non-specificity. Therefore, unacceptable positive results may be caused by biological false positive of ELISA method.
4. In the early stage of infection, antibody did not occur or has a low titer, and these situations will lead to negative results.

