



## User's Manual

# Swine IL-10 ELISA Kit



DEIA-T6106



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

For the quantitative determination of swine interleukin 10 (IL-10) concentrations in cell culture supernates, serum, and plasma. This package insert must be read in its entirety before using this product.

### General Description

Interleukin 10, also known as cytokine synthesis inhibitory factor (CSIF), is the charter member of the IL-10 cytokine family. This family currently comprises IL-10, IL-19, IL-20, IL-22, IL-24, and IL-26/AK155 (1-4). All IL-10 family members are secreted  $\alpha$ -helical proteins. Porcine IL-10 is a secreted, possibly glycosylated, polypeptide with an 18 kDa molecular weight (5). Based on human studies, porcine IL-10 is likely to circulate as a nondisulfide-linked homodimer (3). Porcine IL-10 is synthesized as a 175 amino acid (aa) precursor with an 18 aa signal sequence and a 157 aa mature form. The mature segment has one potential N-linked glycosylation site plus four cysteines which form two intrachain disulfide bridges (6). Mature porcine IL-10 shows 71%, 70%, 76%, 75%, 77%, and 71% aa sequence identity to rat (7), mouse (8), human (9), guinea pig (10), canine (11), and cotton rat (12) IL-10, respectively. Upon activation, mammalian cells known to secrete IL-10 include NK cells (13), cytotoxic CD8<sup>+</sup> T cells secreting Th2-like cytokines (14), CD4<sup>+</sup>CD45RA<sup>+</sup> (memory) Th1 and Th2 cells (15), macrophages (16), monocytes (17), CD5<sup>+</sup> and CD5-B cells (18, 19), dendritic cells (20, 21), hepatic stellate (Ito cells) (22), keratinocytes (23), melanoma cells (24), mast cells (25), placental cytotrophoblasts (26), and fetal erythroblasts (27).


The functional receptor for IL-10 (IL-10 R) in pigs has not been reported. By analogy to human, it would be expected to be composed of two 110 kDa  $\alpha$ -chains (or IL-10 R1) and two 75 kDa  $\beta$ -chains (or IL-10 R2) (28-31). The  $\alpha$ -chain binds IL-10 and transduces a signal in the presence of a  $\beta$ -chain complex (29). Both receptors are members of the class II cytokine receptor family (CRF2) that is characterized by the presence of type III fibronectin domains and conserved tryptophans (29, 31). This class does not possess the WSXWS motif characteristic of the class I CRF. There is no significant sequence identity (< 30%) between human IL-10 R1 and IL-10 R2.

IL-10 has myriad effects on a variety of cell types. On activated B cells, IL-10 can induce plasma cell formation (32) and the secretion of either IgG (33, 34) or IgA (in the presence of TGF- $\beta$  1 and/or IL-4) (34, 35). In the presence of IL-2, CD56<sup>+</sup> NK cells will respond to IL-10 with increased proliferation plus IFN- $\gamma$  and TNF- $\alpha$  secretion (36). Conversely, on macrophages, IL-10 is known to downregulate IL-1, TNF- $\alpha$ , and IL-6 production (37). On dendritic cells, IL-10 has been shown to interfere with antigen-presenting cell function by downmodulating stimulatory and co-stimulatory molecules (38, 39). On monocytes, IL-10 is reported to direct monocyte differentiation into cytotoxic CD16<sup>+</sup> macrophages rather than antigen-presenting dendritic cells (40-42).

### Principles of Testing

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for IL-10 has been pre-coated onto a microplate. Standard, control, or sample and the working solution of Biotin-Conjugate are pipetted into the wells. Following incubation and wash steps, any IL-10 present is bound by the immobilized antibody and the detection antibody specific for IL-10 binds to the combination of capture antibody-IL-10 in sample. Following a wash to remove any unbound combination, and

enzyme conjugate is added to the wells. Following incubation and wash steps a substrate is added. A coloured product is formed in proportion to the amount of IL-10 present in the sample. The reaction is terminated by addition of acid and absorbance is measured at 450nm. A standard curve is prepared from seven IL-10 standard dilutions and IL-10 sample concentration determined.



**Figure 1: Schematic diagram of the assay**

## Reagents And Materials Provided

1. Aluminium pouches with a Microwell Plate coated with antibody to swine IL-10 (8×12)
2. 2 vials swine IL-10 Standard lyophilized, 1000 pg/ml upon reconstitution
3. 2 vials concentrated Biotin-Conjugate anti-swine IL-10 antibody
4. 2 vials Streptavidin-HRP solution,
5. 1 bottle Standard/sample Diluent
6. 1 bottle Biotin-Conjugate antibody Diluent
7. 1 bottle Streptavidin-HRP Diluent
8. 1 bottle Wash Buffer Concentrate 20x (PBS with 1% Tween-20)
9. 1 vial Substrate Solution
10. 1 vial Stop Solution
11. 4 pieces Adhesive Films
12. package insert

NOTE: [96 Tests]

## Materials Required But Not Supplied

1. Microplate reader (450nm).
2. Micro-pipette and tips: 0.5-10, 2-20, 20-200, 200-1000µL.
3. 37°C incubator, double-distilled water or deionized water, coordinate paper, graduated cylinder.

## Storage

For more detailed information, please download the following document on our website.

## Specimen Collection And Preparation

1. **Cell Culture Supernates**-Remove particulates by centrifugation.
2. **Serum**-Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at approximately 1000 x g. Remove serum, avoid hemolysis and high blood lipid samples.
3. **Plasma**-Recommended EDTA as an anticoagulant in plasma. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection.
4. Assay immediately or aliquot and store samples at -20°C. Avoid repeated freeze-thaw cycles.
5. Dilute samples at the appropriate multiple (recommended to do pre-test to determine the dilution factor).

**Note: The normal swine serum or plasma samples are suggested to make a 1:2 dilution.**

## Plate Preparation

1. Prepare all reagents and working standards as directed in the previous sections.
2. Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running blanks and standards. Remove extra microwell strips from holder and store in foil bag with the desiccant provided at 2-8°C sealed tightly.
3. Add 100µL of Standard, control, or sample, per well, then add 50 µL of the working solution of Biotin-Conjugate to each well. Cover with the adhesive strip provided and incubate 2 hours at RT. Adequate mixing is very important for good result. Use a mini-vortexer at the lowest frequency.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (350µL) using a squirt bottle, manifold dispenser or auto-washer. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µL of the working solution of Streptavidin-HRP to each well. Cover with a new adhesive strip and incubate for 30 minutes at RT. Avoid placing the plate in direct light.
6. Repeat the aspiration/wash as in step 4.
7. Add 100 µL of Substrate Solution to each well. Incubate for 10-20 minutes at RT. Avoid placing the plate in direct light.
8. Add 100 µL of Stop Solution to each well. Gently tap the plate to ensure thorough mixing.
9. Determine the optical density of each well immediately, using a microplate reader set to 450 nm. (optionally 630nm as the reference wave length; 610-650nm is acceptable)

## Reagent Preparation

1. Bring all reagents to room temperature before use.
2. **Wash Buffer**-Dilute 10mL of Wash Buffer Concentrate into deionized or distilled water to prepare 200mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
3. **Standard**-Reconstitute the Standard with 1.0mL of Standard/sample Diluent. This reconstitution produces

a stock solution of 1000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.

Pipette 500µL of Standard/sample Diluent into the 1000 pg/mL tube and the remaining tubes. Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 1000 pg/mL standard serves as the high standard. The Standard/sample Diluent serves as the zero standard (0 pg/mL).


If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.

**4. Working solution of Biotin-Conjugate anti-swine IL-10 antibody:** Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody Diluent in a clean plastic tube.

**The working solution should be used within one day after dilution.**

**5. Working solution of Streptavidin-HRP:** Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP Diluent in a clean plastic tube.

**The working solution should be used within one day after dilution.**



**Figure 2: Preparation of IL-10 standard dilutions**

## Assay Procedure

### Assay procedure summary

1. Prepare all reagents and standards as directed.
2. Add 100µL standard or samples to each well, and add 50µL working solution of Biotin-Conjugate antibody to each well, incubate 120 minutes, RT.
3. Aspirate and wash 4 times.
4. Add 100µL working solution of Streptavidin-HRP to each well, incubate 30 minutes, RT.
5. Aspirate and wash 4 times.
6. Add 100µL Substrate solution to each well, incubate 10-20 minutes, RT. Protect from light.
7. Add 100µL Stop solution to each well. Read at 450nm within 30 minutes.

## Calculation

1. Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density.
2. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean


absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.

- The data may be linearized by plotting the log of the IL-10 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

## Typical Standard Curve

Typical data using the IL-10 ELISA (Measuring wavelength: 450nm, Reference wavelength: 630nm)

Standardized (pg/ml)	OD.	OD.	Average	Corrected
0	0.041	0.043	0.042	---
15.625	0.073	0.069	0.071	0.076
31.25	0.116	0.114	0.115	0.125
62.5	0.208	0.201	0.205	0.221
125	0.404	0.406	0.405	0.401
250	0.754	0.750	0.752	0.724
500	1.234	1.229	1.232	1.254
1000	2.014	2.009	2.012	2.007



Representative standard curve for IL-10 ELISA. IL-10 was diluted in serial two-fold steps in Sample Diluent.

**Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.**

## Performance Characteristics

- When mixing or reconstituting protein solutions, always avoid foaming.
- To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

3. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
4. Substrate Solution should remain colorless until added to the plate. Stop Solution should be added to the plate in the same order as the Substrate Solution. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
5. A standard curve should be generated for each set of samples assayed. According to the content of tested factors in the sample, appropriate diluted or concentrated samples, it is best to do pre-experiment.
1. REPEATABILITY: The coefficient of variation of both intra-assay and inter-assay were less than 10%.
2. SENSITIVITY: The minimum detectable dose was 7pg/mL.
3. SPECIFICITY: This assay recognizes both recombinant and natural porcine IL-10. The factors listed below were prepared at 50 ng/mL in Calibrator Diluent RD6-33 and assayed for cross-reactivity. Preparations of the following factors at the same concentrations in a mid-range porcine IL-10 control were assayed for interference. No significant cross-reactivity or interference was observed.

Factors assayed for cross-reactivity

Recombinant human	Recombinant mouse	Recombinant rat
IL-4	IL-4	IL-2_
IFN- $\gamma$	IFN- $\gamma$	IL-4_
IL-10	IL-10	IFN- $\gamma$
sIL-10R	sIL-10R	TNF- $\alpha$

## Precautions

1. Store kit reagents between 2°C and 8°C. After use all reagents should be immediately returned to cold storage(2°C to 8°C).
2. Please perform simple centrifugation to collect the liquid before use.
3. To avoid cross contamination, please use disposable pipette tips.
4. The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material. Avoid contact of skin or mucous membranes with kit reagents or specimens. In the case of contact with skin or eyes wash immediately with water.
5. Use clean, dedicated reagent trays for dispensing the washing liquid, conjugate and substrate reagent. Mix all reagents and samples well before use.
6. After washing microtiter plate should be fully pat dried. Do not use absorbent paper directly into the enzyme reaction wells.
7. Do not mix or substitute reagents with those from other lots or other sources. Do not use kit reagents beyond expiration date on label.
8. Each sample, standard, blank and optional control samples should be assayed in duplicate or triplicate.
9. Adequate mixing is very important for good result. Use a mini-vortexer at the lowest frequency or Shake by hand at 10min interval when there is no vortexer.
10. Avoid microtiter plates drying during the operation.
11. Dilute samples at the appropriate multiple, and make the sample values fall within the standard curve. If samples generate values higher than the highest standard, dilute the samples and repeat the assay.
12. Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time and temperature, and kit age can cause variation in binding.
13. This method can effectively eliminate the interference of the soluble receptors, binding proteins and other



factors in biological samples.

## References

1. Rich, B.E. and T.S. Kupper (2001) *Curr. Biol.* 11:R531.
2. Gruenberg, B.H. et al. (2001) *Genes Immun.* 2:329.
3. Moore, K.W. et al. (2001) *Annu. Rev. Immunol.* 19:683.
4. Fickenscher, H. et. al. (2002) *Trends Immunol.* 23:89.
5. Blacho, G. et al. (1995) *Proc. Natl. Acad. Sci. USA* 92:2800.
6. Windsor, W.T. et al. (1993) *Biochemistry* 32:8807.
7. Feng, L. et al. (1993) *Biochem. Biophys. Res. Commun.* 192:452.
8. Moore, K.W. et al. (1990) *Science* 248:1230.
9. Vieira, P. et al. (1991) *Proc. Natl. Acad. Sci. USA* 88:1172.
10. Scarozza, A.M. et al. (1998) *Cytokine* 10:851.
11. Lu, P. et al. (1995) *J. Interf. Cytokine Res.* 15:1103.
12. Langley, R.J. et al. (2001) GenBank Accession #AAK04013.
13. Mehrotra, P.T. et al. (1998) *J. Immunol.* 160:2637.
14. Sad, S. et al. (1995) *Immunity* 2:271.
15. Yssel, H. et al. (1992) *J. Immunol.* 149:2378.
16. Panuska, J.R. et al. (1995) *J. Clin. Invest.* 96:2445.
17. Hodge, S. et al. (1999) *Scand. J. Immunol.* 49:548.
18. Spencer, N.F.L. and R.A. Daynes (1997) *Int. Immunol.* 9:745.
19. O'Garra, A. et al. (1990) *Int. Immunol.* 2:821.
20. Iwasaki, A. and B.L. Kelsall (1999) *J. Exp. Med.* 190:229.
21. Rea, D. et al. (2000) *Blood* 95:3162.
22. Wang, S.C. et al. (1998) *J. Biol. Chem.* 273:302.
23. Grewe, M. et al. (1995) *J. Invest. Dermatol.* 104:3.
24. Sato, T. et al. (1996) *Clin. Cancer Res.* 2:1383.
25. Ishizuka, T. et al. (1999) *Clin. Exp. Allergy* 29:1424.
26. Roth, I. et al. (1996) *J. Exp. Med.* 184:539.
27. Sennikov, S.V. et al. (2002) *Cytokine* 17:221.
28. Kotenko, S.V. et al. (1997) *EMBO J.* 16:5894.
29. Lutfalla, G. et al. (1993) *Genomics* 16:366.
30. Gibbs, V.C. and D. Pennica (1997) *Gene* 186:97.
31. Liu, Y. et al. (1994) *J. Immunol.* 152:1821.
32. Rousset, F. et al. (1995) *Int. Immunol.* 7:1243.
33. Briere, F. et al. (1994) *J. Exp. Med.* 179:757.
34. Defrance, T. et al. (1992) *J. Exp. Med.* 175:671.





- 35. Marconi, M. et al. (1998) Clin. Exp. Immunol. 112:528.
- 36. Carson, W.E. et al. (1995) Blood 85:3577.
- 37. Fiorentino, D.F. et al. (1991) J. Immunol. 147:3815.
- 38. Qzawa, H. et al. (1996) Eur. J. Immunol. 26:648.
- 39. Buelens, C. et al. (1995) Eur. J. Immunol. 25:2668.
- 40. Olikowsky, T. et al. (1997) Immunology 91:104.
- 41. Calzada-Wack, J.C. et al. (1996) J. Inflamm. 46:78.
- 42. Allavena, P. et al. (1998) Eur. J. Immunol. 28:359.
- 43. Bradford, M.M. (1976) Anal. Biochem. 72:248.

