



## User's Manual

# Human Osteocalcin (1-43/49) ELISA kit



DEIA-XY10



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

This ELISA (enzyme-linked immunosorbent assay) kit is intended for measurement of both human osteocalcin (1-49) and osteocalcin (1-43) (also referred as N-terminal & mid-regional osteocalcin) levels in test samples. This test is useful for assessing the bone formation activity or osteoblast activity in individuals associated with changes in the rate of bone turnover in metabolic bone disease, such as osteoporosis, primary hyperparathyroidism, hyperthyroidism, Paget's disease, and renal osteodystrophy.

### Storage

This test kit must be stored at 2°C-8°C upon receipt. For the expiration date of the kit refer to the label on the kit box.

All components are stable until this expiration date.

