



User's Manual

TNF α -Blocker monitoring, infliximab drug level ELISA Kit



DEIA024J



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: info@creative-diagnostics.com  Web: www.creative-diagnostics.com

PRODUCT INFORMATION

Intended Use

The assay is a Enzyme Immuno assay intended for the quantitative determination of free therapeutic TNF α antibodies Infliximab in EDTA plasma and serum.

Storage

Unopened Kit: Store at 2-8°C. Do not use past kit expiration date.

Precision

Intra-Assay Precision: 1.8%-9.7%

Inter-Assay Precision: 4%-11%

Sensitivity

2 ng/mL