



User's Manual

Taenia Solium IgG ELISA Kit



DEIA1793



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: info@creative-diagnostics.com  Web: www.creative-diagnostics.com

PRODUCT INFORMATION

Intended Use

For the qualitative screening of serum IgG antibodies to Taenia solium using an Enzyme Linked Immunoabsorbant Assay (ELISA) technique.

Principles of Testing

The micro test wells are coated with T. solium cyst fluid antigen. During the first incubation with the diluted sera, any antibodies which are reactive with the antigen will bind to the coated wells. After washing to remove the rest of the sample, the Enzyme Conjugate is added. If antibodies have been bound to the wells, the Enzyme Conjugate will then bind to these antibodies. After another series of washes, a chromogen (tetramethylbenzidine or TMB) is added. If the Enzyme Conjugate is present, the peroxidase will catalyze a reaction that consumes the peroxide and turns the chromogen from clear to blue. Addition of the Stop Solution ends the reaction and turns the blue color to a bright yellow color. The reaction may then be read visually or with an ELISA reader.

Reagents And Materials Provided

Item	Description	Symbol
Test Strips	Microwells containing T. solium antigens - 96 test wells in a test strip holder.	MT PLATE
Enzyme Conjugate	One (1) bottle containing 11 ml of Protein A conjugated to peroxidase.	CONJ
Positive Control	One (1) vial containing 1 ml of diluted positive rabbit serum.	CONTROL +
Negative Control	One (1) vial containing 1 ml of diluted negative human serum.	CONTROL -
Chromogen	One (1) bottle containing 11 ml of the chromogen tetramethylbenzidine (TMB).	SUBS TMB
Wash Concentrate (20X)	One (1) bottle containing 25 ml of concentrated buffer and surfactant.	WASH BUF
Dilution Buffer	Two (2) bottles containing 30 ml of buffered protein solution.	SPECIM DIL
Stop Solution	One (1) bottle containing 11 ml of 0.73 M phosphoric acid.	SOLN

Materials Required But Not Supplied

1. Pipettes
2. Squeeze bottle for washing strips (narrow tip is recommended)
3. Reagent grade water and graduated cylinder
4. Tubes for sample dilution
5. Absorbent paper
6. ELISA plate reader with a 450 nm and a 650 to 620 nm filter (optional if results are read visually)

Storage

Reagents, strips and bottled components: Store between 2-8°C.

Squeeze bottle containing diluted wash buffer may be stored at room temperature.

Specimen Collection And Preparation

Serum

Collection and Preparation of Serum Serological specimens should be collected under aseptic conditions. Coagulate blood and remove serum. Hemolysis is avoided through prompt separation of the serum from the clot. Serum should be stored at 2-8°C if it is to be analyzed within a few days. Serum may be held for 3 to 6 months by storage at -20°C or lower. Lipemic and strongly hemolytic serum should be avoided. Do not heat inactivate serum and avoid repeated freezing and thawing of samples.

Test samples

Make a 1:64 dilution of sera using the dilution buffer (e.g. 5 µl sera and 315 µl dilution buffer).

Reagent Preparation

Wash Buffer

Remove cap and add contents of bottle to 475 ml of reagent grade water. Place diluted wash buffer into a squeeze bottle with a narrow tip opening.

Note:

- 1) Washings consist of filling to the top of each well, shaking out the contents and refilling.
- 2) Avoid generating bubbles in the wells during the washing steps.

Materials Provided

Cysticercosis Serology Microwell ELISA Kit

Assay Procedure

Performance of Test

1. Break off number of wells needed (two for controls plus number of samples) and place in strip holder.
2. Add 100 µl (or two drops) of the negative control to well #1, 100 µl of the positive control to well #2 and 100 µl of the diluted (1:64) test samples to the remaining wells.

Note: Negative and positive controls are supplied prediluted. Do not dilute further.

3. Incubate at room temperature (15 to 25 °C) for 10 minutes.
4. Shake out contents and wash 3 times with the diluted wash buffer.
5. Add 2 drops of Enzyme Conjugate to each well.
6. Incubate at room temperature for 5 minutes.
7. Shake out contents and wash 3 times with wash buffer. Slap wells against paper towels to remove excess moisture.
8. Add 2 drops of the Chromogen to every well.
9. Incubate at room temperature for 5 minutes.
10. Add 2 drops of the Stop Solution and mix by tapping strip holder.

Reading of Results

Visually: Look at each well against a white background (e.g. paper towel) and record as clear or +, ++ or +++ reaction.

ELISA Reader: Zero reader on air. Set for bichromatic readings at 450/650-620 nm.

Quality Control

The use of controls allows validation of kit stability. The kit should not be used if any of the controls are out of range.

Expected values for the controls are:

Negative - 0.0 to 0.3 OD units

Positive - 0.5 OD units and above

Interpretation Of Results

Interpretation of Results - ELISA Reader

Zero ELISA reader on air. Read all wells at 450/650 to 620 nm.

Positive - Absorbance reading greater than 0.3 OD units.

Negative - Absorbance reading less than 0.3 OD units.

Interpretation of Results -Visual

Compare results to the controls. A sample should be interpreted as positive if the degree of color development is obvious and significant.

Performance Characteristics

	Reference Method	
	+	-
+	72	2
-	10	46

Specificity

Forty-eight normal samples were tested in the ELISA kit; 46/48 samples were negative in the ELISA kit giving a specificity of 96%.

Precautions

1. Do not use solutions if they precipitate or become cloudy. Wash concentrate may show crystallization upon

storage at 2-8°C. Crystallization will disappear after dilution to working strength.

2. Do not use serum that may have supported microbial growth, or is cloudy due to high lipid content. Samples high in lipids should be clarified before use.
3. Treat all sera as if capable of being infectious. Negative control has been tested and found negative for Hepatitis B surface antigen and for the antibody to HIV by required test methods. This product should be used under appropriate safety conditions that would be used for any potentially infectious agent.
4. Do not add azides to the samples or any of the reagents.
5. Negative control has excessive color after development. Reason: inadequate washings. Correction: wash more vigorously. Remove excessive liquid from the wells by tapping against an absorbent towel. Do not allow test wells to dry out.

