



## User's Manual

# Human Influenza B HA ELISA Kit



DEIA249



5 plates



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

The influenza B HA ELISA Pair Set is for the quantitative determination of influenza B HA.

This ELISA Pair Set contains the basic components required for the development of sandwich ELISAs.

### General Description

The influenza viral Hemagglutinin (HA) protein is a homo trimer with a receptor binding pocket on the globular head of each monomer. HA has at least 18 different antigens. These subtypes are named H1 through H18. HA has two functions. Firstly, it allows the recognition of target vertebrate cells, accomplished through the binding to these cells' sialic acid-containing receptors. Secondly, once bound it facilitates the entry of the viral genome into the target cells by causing the fusion of host endosomal membrane with the viral membrane. The influenza virus Hemagglutinin (HA) protein is translated in cells as a single protein, HA0, or hemagglutinin precursor protein. For viral activation, hemagglutinin precursor protein (HA0) must be cleaved by a trypsin-like serine endoprotease at a specific site, normally coded for by a single basic amino acid (usually arginine) between the HA1 and HA2 domains of the protein. After cleavage, the two disulfide-bonded protein domains produce the mature form of the protein subunits as a prerequisite for the conformational change necessary for fusion and hence viral infectivity.

### Principles of Testing

The CD ELISA Pair Set is a solid phase sandwich ELISA (Enzyme-Linked Immunosorbent Assay). It utilizes a monoclonal antibody specific for Influenza B HA coated on a 96-well plate. Standard sand samples are added to the wells, and any Influenza B HA present binds to the immobilized antibody. The wells are washed and a horseradish peroxidase conjugated rabbit anti-Influenza B HA poly clonal antibody is then added, producing an antibody-antigen-antibody "sandwich". The wells are again washed and TMB substrate solution is loaded, which produces color in proportion to the amount of Influenza B HA present in the sample. To end the enzyme reaction, the stop solution is added and absorbances of the microwell are read at 450 nm.

### Reagents And Materials Provided

**Bring all reagents to room temperature before use.**

**Capture Antibody:** 0.5 mg/mL of mouse anti-influenza B HA monoclonal antibody (in PBS, pH 7.4). Dilute to a working concentration of 2 µg/mL in PBS before coating.

**Detection Antibody:** 0.2 mg/mL rabbit anti-influenza B HA polyclonal antibody conjugated to horseradish-peroxidase (HRP) (in PBS, 50 % HRP-Protector, pH 7.4). Dilute to working concentration of 0.5 µg/mL in detection antibody dilution buffer before use.

**Standard:** Each vial contains 145 ng of recombinant influenza B HA. Reconstitute with 1 mL detection antibody dilution buffer. After reconstitution, store at -20°C to -80°C in a manual defrost freezer. A seven-point standard curve using 2-fold serial dilutions in sample dilution buffer, and a high standard of 4 ng/mL is recommended.

## Materials Required But Not Supplied

**PBS:** 136.9 mM NaCl, 10.1 mM Na<sub>2</sub>HPO<sub>4</sub>, 2.7 mM KCl, 1.8 mM KH<sub>2</sub>PO<sub>4</sub>, pH 7.4, 0.2 µm filtered

**TBS:** 20 mM Tris, 150 mM NaCl, pH 7.4

**Wash Buffer:** 0.05% Tween20 in TBS, pH 7.2-7.4

**Blocking Buffer:** 2% BSA in Wash Buffer

**Sample dilution buffer:** 0.1% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

**Detection antibody dilution buffer:** 0.5% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

**Substrate Solution:** To achieve best assay results, fresh substrate solution is recommended

**Substrate stock solution:** 10mg / ml TMB (Tetramethylbenzidine) in DMSO

**Substrate dilution buffer:** 0.05M Na<sub>2</sub>HPO<sub>4</sub> and 0.025M citric acid; adjust pH to 5.5

**Substrate working solution:** For each plate dilute 250 µl substrate stock solution in 25 ml substrate dilution buffer and then add 80 µl 0.75% H<sub>2</sub>O<sub>2</sub>, mix it well

**Stop Solution:** 2 N H<sub>2</sub>SO<sub>4</sub>

## Storage

**Capture Antibody:** Aliquot and store at -20°C to -80°C for up to 6 months from date of receipt. Avoid repeated freeze-thaw cycles.

**Detection Antibody:** Store at 4°C and protect it from prolonged exposure to light for up to 6 months from date of receipt. Do not freeze!

**Standard:** Store lyophilized standard at -20°C to -80°C for up to 6 months from date of receipt. Aliquot and store the reconstituted standard at -80°C for up to 1 month. Avoid repeated freezethaw cycles.

## Plate Preparation

1. Dilute the capture antibody to the working concentration in PBS. Immediately coat a 96-well microplate with 100µL per well of the diluted capture antibody. Seal the plate and incubate overnight at 4°C.
2. Aspirate each well and wash with at least 300µl wash buffer, repeating the process two times for a total of three washes. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining wash buffer by inverting the plate and blotting it against clean paper towels.
3. Block plates by adding 300 µL of blocking buffer to each well. Incubate at room temperature for a minimum of 1 hour.
4. Repeat the aspiration/wash as in step 2. The plates are now ready for sample addition.

## Assay Procedure

1. Add 100 µL of sample or standards in sample dilution buffer per well. Seal the plate and incubate 2 hours at room temperature.
2. Repeat the aspiration/wash as in step 2 of plate preparation.

3. Add 100  $\mu$ L of the detection antibody, diluted in antibody dilution buffer, to each well. Seal the plate and incubate 1 hour at room temperature.
4. Repeat the aspiration/wash as in step 2 of plate preparation.
5. Add 200  $\mu$ L of substrate solution to each well. Incubate for 20 minutes at room temperature (if substrate solution is not as requested, the incubation time should be optimized). Avoid placing the plate in direct light.
6. Add 50  $\mu$ L of stop solution to each well. Gently tap the plate to ensure thorough mixing.
7. Determine the optical density of each well immediately, using a microplate reader set to 450 nm.


## Calculation

1. Calculate the mean absorbance for each set of duplicate standards, controls and samples. Subtract the mean zero standard absorbance from each.
2. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
3. To determine the concentration of the unknowns, find the unknowns' mean absorbance value on the y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the x-axis and read the concentration. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
4. Alternatively, computer-based curve-fitting statistical software may also be employed to calculate the concentration of the sample.

## Typical Standard Curve

This standard curve is only for demonstration purposes. A standard curve should be generated for each assay.

Concentration (pg/ml)	Zero standard subtracted OD
<b>0</b>	<b>0.000</b>
<b>62.5</b>	<b>0.034</b>
<b>125</b>	<b>0.065</b>
<b>250</b>	<b>0.137</b>
<b>500</b>	<b>0.257</b>
<b>1000</b>	<b>0.512</b>
<b>2000</b>	<b>1.007</b>
<b>4000</b>	<b>2.089</b>




## Sensitivity

The minimum detectable dose of Influenza B HA was determined to be approximately **62.5 pg/ml**.

This is defined as at least three times standard deviations above the mean optical density of 10 replicates of the zero standard.

## Precautions

The Stop Solution suggested for use with this Pair Set is an acid solution. Wear eye, hand, face, and clothing protection when using this material.



## Trouble Shooting

Problems	Possible Sources	Solutions
<b>No signal</b>	Incorrect or no Detection Antibody was added	Add appropriate Detection Antibody and continue
	Substrate solution was not added	Add substrate solution and continue
	Incorrect storage condition	Check if the kit is stored at recommended condition and used before expiration date
<b>Poor Standard Curve</b>	Standard was incompletely reconstituted or was inappropriately stored	Aliquot reconstituted standard and store at -80 °C
	Imprecise / inaccurate pipetting	Check / calibrate pipettes
	Incubations done at inappropriate temperature, timing or agitation	Follow the general ELISA protocol
	Background wells were contaminated	Avoid cross contamination by using the sealer appropriately
<b>Poor detection value</b>	The concentration of antigen in samples was too low	Enriching samples to increase the concentration of antigen
	Samples were ineffective	Check if the samples are stored at cold environment. Detect samples in timely manner
<b>High Background</b>	Insufficient washes	Use multichannel pipettes without touching the reagents on the plate
		Increase cycles of washes and soaking time between washes
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells
	Materials were contaminated.	Use clean plates, tubes and pipettes tips
<b>Non-specificity</b>	Samples were contaminated	Avoid cross contamination of samples
	The concentration of samples was too high	Try higher dilution rate of samples

