



User's Manual

Human Interleukin 1 Receptor-Like 1, IL1RL1 ELISA Kit



DEIA260



5 plates



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 **Address: 45-1 Ramsey Road, Shirley, NY 11967, USA**

 **Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)**  **Fax: 1-631-938-8221**

 **Email: info@creative-diagnostics.com**  **Web: www.creative-diagnostics.com**

PRODUCT INFORMATION

Intended Use

The Human IL1RL1 / DER4 ELISA Pair Set is for the quantitative determination of Human IL1RL1/DER5.

This ELISA Pair Set contains the basic components required for the development of sandwich ELISAs.

General Description

Interleukin-1 receptor-like 1, also known as protein ST2, IL1RL1 and DER4, is a single-pass type I membrane and a secreted protein which belongs to the interleukin-1 receptor family. IL1RL1 / DER4 has a soluble secreted form (sST2) and a transmembrane form (ST2L). IL1RL1 / DER4 contains three Ig-like C2-type (immunoglobulin-like) domains and one TIR domain. IL1RL1 / DER4 is highly expressed in kidney, lung, placenta, stomach, skeletal muscle, colon and small intestine. Isoform A of IL1RL1 is prevalently expressed in the lung, testis, placenta, stomach and colon. Isoform B of IL1RL1 is more abundant in the brain, kidney and the liver. Isoform C of IL1RL1 is not detected in brain, heart, liver, kidney and skeletal muscle. The coding gene forms a cytokine receptor gene cluster together with IL1R1(interleukin 1 receptor, type I), IL1R2 (interleukin 1 receptor, type II) and IL1RL2 (interleukin 1 receptor-like 2) genes in a region mapped to chromosome 2q12. As an important mediator involved in many immune and inflammatory responses, this cytokine has been implicated as a regulator of both the development and effector phases of type 2 helper T cell responses, and as a negative feedback modulator of macrophage pro-inflammatory function. IL33 is a specific ligand of the protein IL1RL1 and induces production of Th2 cytokines.

Principles of Testing

The CD ELISA Pair Set is a solid phase sandwich ELISA (Enzyme-Linked Immunosorbent Assay). It utilizes a monoclonal antibody specific for Human IL1RL1 / DER4 coated on a 96-well plate. Standards and samples are added to the wells, and any Human IL1RL1 / DER4 present binds to the immobilized antibody. The wells are washed and a horseradish peroxidase conjugated mouse antiHuman IL1RL1 / DER4 monoclonal antibody is then added, producing an antibody-antigen-antibody "sandwich". The wells are again washed and TMB substrate solution is loaded, which produces color in proportion to the amount of Human IL1RL1 / DER4 present in the sample. To end the enzyme reaction, the stop solution is added and absorbances of the microwell are read at 450 nm.

Reagents And Materials Provided

Bring all reagents to room temperature before use.

Capture Antibody: 1 mg/mL of mouse anti-Human IL1RL1 / DER4 monoclonal antibody (in PBS, pH 7.4). Dilute to a working concentration of 2 µg/mL in PBS before coating.

Detection Antibody: 0.2 mg/mL of mouse anti-Human IL1RL1 / DER4 monoclonal antibody conjugated to horseradish-peroxidase (HRP) (in PBS, 50 % HRP-Protector, pH 7.4, store at 4°C). Dilute to working concentration of 1.2 µg/mL in detection antibody dilution buffer before use.

Standard: Each vial contains 28 ng of recombinant Human IL1RL1 / DER4. Reconstitute with 1 mL detection antibody dilution buffer. After reconstitution, store at -20°C to -80°C in a manual defrost freezer. A seven-

point standard curve using 2-fold serial dilutions in sample dilution buffer, and a high standard of 500 pg/mL is recommended.

Materials Required But Not Supplied

PBS: 136.9 mM NaCl, 10.1 mM Na₂HPO₄, 2.7 mM KCl, 1.8 mM KH₂PO₄, pH 7.4, 0.2 µm filtered

TBS: 20 mM Tris, 150 mM NaCl, pH 7.4

Wash Buffer: 0.05% Tween20 in TBS, pH 7.2-7.4

Blocking Buffer: 2% BSA in Wash Buffer

Sample dilution buffer: 0.1% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

Detection antibody dilution buffer: 0.5% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

Substrate Solution: To achieve best assay results, fresh substrate solution is recommended

Substrate stock solution: 10mg / ml TMB (Tetramethylbenzidine) in DMSO

Substrate dilution buffer: 0.05M Na₂HPO₄ and 0.025M citric acid; adjust pH to 5.5

Substrate working solution: For each plate dilute 250 µl substrate stock solution in 25 ml substrate dilution buffer and then add 80 µl 0.75% H₂O₂, mix it well

Stop Solution: 2 N H₂SO₄

Storage

Capture Antibody: Aliquot and store at -20°C to -80°C for up to 6 months from date of receipt. Avoid repeated freeze-thaw cycles.

Detection Antibody: Store at 4°C and protect it from prolonged exposure to light for up to 6 months from date of receipt. Do not freeze!

Standard: Store lyophilized standard at -20°C to -80°C for up to 6 months from date of receipt. Aliquot and store the reconstituted standard at -80°C for up to 1 month. Avoid repeated freezethaw cycles.

Plate Preparation

1. Dilute the capture antibody to the working concentration in PBS. Immediately coat a 96-well microplate with 100µL per well of the diluted capture antibody. Seal the plate and incubate overnight at 4°C.
2. Aspirate each well and wash with at least 300µl wash buffer, repeating the process two times for a total of three washes. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining wash buffer by inverting the plate and blotting it against clean paper towels.
3. Block plates by adding 300 µL of blocking buffer to each well. Incubate at room temperature for a minimum of 1 hour.
4. Repeat the aspiration/wash as in step 2. The plates are now ready for sample addition.

Assay Procedure

1. Add 100 μ L of sample or standards in sample dilution buffer per well. Seal the plate and incubate 2 hours at room temperature.
2. Repeat the aspiration/wash as in step 2 of plate preparation.
3. Add 100 μ L of the detection antibody, diluted in antibody dilution buffer, to each well. Seal the plate and incubate 1 hour at room temperature.
4. Repeat the aspiration/wash as in step 2 of plate preparation.
5. Add 200 μ L of substrate solution to each well. Incubate for 20 minutes at room temperature (if substrate solution is not as requested, the incubation time should be optimized). Avoid placing the plate in direct light.
6. Add 50 μ L of stop solution to each well. Gently tap the plate to ensure thorough mixing.
7. Determine the optical density of each well immediately, using a microplate reader set to 450 nm.


Calculation

1. Calculate the mean absorbance for each set of duplicate standards, controls and samples. Subtract the mean zero standard absorbance from each.
2. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
3. To determine the concentration of the unknowns, find the unknowns' mean absorbance value on the y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the x-axis and read the concentration. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
4. Alternatively, computer-based curve-fitting statistical software may also be employed to calculate the concentration of the sample.

Typical Standard Curve

This standard curve is only for demonstration purposes. A standard curve should be generated for each assay.

Concentration (pg/mL)	Zero standard subtracted OD
0	0
7.81	0.044
15.63	0.085
31.25	0.169
62.5	0.338
125	0.684
250	1.296
500	2.251




Sensitivity

The minimum detectable dose of Human IL1RL1 / DER4 was determined to be approximately **7.8 pg/ml**.

This is defined as at least three times standard deviations above the mean optical density of 10 replicates of the zero standard.

Precautions

The Stop Solution suggested for use with this Pair Set is an acid solution. Wear eye, hand, face, and clothing protection when using this material.



Trouble Shooting

Problems	Possible Sources	Solutions
No signal	Incorrect or no Detection Antibody was added	Add appropriate Detection Antibody and continue
	Substrate solution was not added	Add substrate solution and continue
	Incorrect storage condition	Check if the kit is stored at recommended condition and used before expiration date
Poor Standard Curve	Standard was incompletely reconstituted or was inappropriately stored	Aliquot reconstituted standard and store at -80 °C
	Imprecise / inaccurate pipetting	Check / calibrate pipettes
	Incubations done at inappropriate temperature, timing or agitation	Follow the general ELISA protocol
	Background wells were contaminated	Avoid cross contamination by using the sealer appropriately
Poor detection value	The concentration of antigen in samples was too low	Enriching samples to increase the concentration of antigen
	Samples were ineffective	Check if the samples are stored at cold environment. Detect samples in timely manner
High Background	Insufficient washes	Use multichannel pipettes without touching the reagents on the plate
		Increase cycles of washes and soaking time between washes
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells
	Materials were contaminated.	Use clean plates, tubes and pipettes tips
Non-specificity	Samples were contaminated	Avoid cross contamination of samples
	The concentration of samples was too high	Try higher dilution rate of samples

