

**User's Manual**

Human Interleukin 20 Receptor, Alpha, IL20RA ELISA Kit

REF DEIA261**5 plates****RUO**

This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics **Address:** 45-1 Ramsey Road, Shirley, NY 11967, USA **Tel:** 1-631-624-4882 (USA) 44-161-818-6441 (Europe) **Fax:** 1-631-938-8221 **Email:** info@creative-diagnostics.com **Web:** www.creative-diagnostics.com

PRODUCT INFORMATION

Intended Use

The human IL20RA / IL20R1 ELISA Pair Set is for the quantitative determination of human IL20RA / IL20R1.

This ELISA Pair Set contains the basic components required for the development of sandwich ELISAs.

General Description

Interleukin-20 receptor subunit alpha also known as IL-20 receptor subunit alpha, IL-20R-alpha, Cytokine receptor family 2 member 8, IL20RA and IL-20R1 is a single-pass type I membrane protein which belongs to the type II cytokine receptor family. IL20RA / IL20R1 contains two fibronectin type-III domains. The interleukin-20-receptor I complex (IL-20-RI) is composed of two chains, IL20RA and IL20RB. IL20RA / IL20R1 forms heterodimer with IL20RB. Its ligands are the three members of the IL19 subfamily of cytokines, IL-19, IL-20 and IL-24. These cytokines are important in the manifestation of psoriatic lesions. All three are capable of signaling through IL20RA / IL20RB complex. The ligand binding to receptor B creating a high-affinity binding site for the receptor A which is recruited to complete the complex. In addition, IL20RA / IL20R1 also forms a heterodimer with the unique and specific receptor IL10RB and functions as the receptor for IL26. IL20RA / IL20R1 is widely expressed with highest levels in skin, testis and brain. The expression of both IL20RA and IL20RB is found to be upregulated in psoriatic skin lesions on keratinocytes.

Principles of Testing

The CD ELISA Pair Set is a solid phase sandwich ELISA (Enzyme-Linked Immunosorbent Assay). It utilizes a monoclonal antibody specific for IL20RA / IL20R1 coated on a 96-well plate. Standards and samples are added to the wells, and any IL20RA / IL20R1 present binds to the immobilized antibody. The wells are washed and a horseradish peroxidase conjugated rabbit anti-IL20RA polyclonal antibody is then added, producing an antibody-antigen-antibody "sandwich". The wells are again washed and TMB substrate solution is loaded, which produces color in proportion to the amount of IL20RA / IL20R1 present in the sample. To end the enzyme reaction, the stop solution is added and absorbances of the microwell are read at 450 nm.

Reagents And Materials Provided

Bring all reagents to room temperature before use.

Capture Antibody: 0.5 mg/mL of mouse anti-IL20RA monoclonal antibody. Dilute to a working concentration of 1 µg/mL in CBS before coating.

Detection Antibody: 0.5 mg/mL rabbit anti-IL20RA polyclonal antibody conjugated to horseradish-peroxidase (HRP). Dilute to working concentration of 0.5 µg/mL in detection antibody diluteion buffer before use.

Standard: Each vial contains 70ng of recombinant IL20RA. Reconstitute with 1 mL detection antibody dilution buffer. After reconstitution, store at -20°C to -80°C in a manual defrost freezer. A seven-point standard curve using 2-fold serial dilutions in sample dilution buffer, and a high standard of 2 ng/mL is recommended.

Materials Required But Not Supplied

CBS: 0.05M Na₂CO₃, NaHCO₃, pH 9.6, 0.2 µm filtered

TBS: 25mM Tris, adjust pH to 7.4 by HCl

Wash Buffer: 0.05% Tween20 in TBS, pH 7.2-7.4

Blocking Buffer: 2% BSA in Wash Buffer

Sample dilution buffer: 0.1% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

Detection antibody dilution buffer: 0.5% BSA in wash buffer, pH 7.2-7.4, 0.2 µm filtered

Substrate Solution: To achieve best assay results, fresh substrate solution is recommended

Substrate stock solution: 10mg / ml TMB (Tetramethylbenzidine) in DMSO

Substrate dilution buffer: 0.05M Na₂HPO₄ and 0.025M citric acid; adjust pH to 5.5

Substrate working solution: For each plate dilute 250 µl substrate stock solution in 25 ml substrate dilution buffer and then add 80 µl 0.75% H₂O₂, mix it well

Stop Solution: 2 N H₂SO₄

Storage

Capture Antibody: Aliquot and store at -20°C to -80°C for up to 6 months from date of receipt. Avoid repeated freeze-thaw cycles.

Detection Antibody: Store at 4°C and protect it from prolonged exposure to light for up to 6 months from date of receipt. Do not freeze!

Standard: Store lyophilized standard at -20°C to -80°C for up to 6 months from date of receipt. Aliquot and store the reconstituted standard at -80°C for up to 1 month. Avoid repeated freezethaw cycles.

Plate Preparation

1. Dilute the capture antibody to the working concentration in CBS. Immediately coat a 96-well microplate with 100µL per well of the diluted capture antibody. Seal the plate and incubate overnight at 4°C.
2. Aspirate each well and wash with at least 300µl wash buffer, repeating the process two times for a total of three washes. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining wash buffer by inverting the plate and blotting it against clean paper towels.
3. Block plates by adding 300 µL of blocking buffer to each well. Incubate at room temperature for a minimum of 1 hour.
4. Repeat the aspiration/wash as in step 2. The plates are now ready for sample addition.

Assay Procedure

1. Add 100 µL of sample or standards in sample dilution buffer per well. Seal the plate and incubate 2 hours at room temperature.
2. Repeat the aspiration/wash as in step 2 of plate preparation.

3. Add 100 μ L of the detection antibody, diluted in antibody dilution buffer, to each well. Seal the plate and incubate 1 hour at room temperature.
4. Repeat the aspiration/wash as in step 2 of plate preparation.
5. Add 200 μ L of substrate solution to each well. Incubate for 20 minutes at room temperature (if substrate solution is not as requested, the incubation time should be optimized). Avoid placing the plate in direct light.
6. Add 50 μ L of stop solution to each well. Gently tap the plate to ensure thorough mixing.
7. Determine the optical density of each well immediately, using a microplate reader set to 450 nm.


Calculation

1. Calculate the mean absorbance for each set of duplicate standards, controls and samples. Subtract the mean zero standard absorbance from each.
2. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
3. To determine the concentration of the unknowns, find the unknowns' mean absorbance value on the y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the x-axis and read the concentration. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
4. Alternatively, computer-based curve-fitting statistical software may also be employed to calculate the concentration of the sample.

Typical Standard Curve

This standard curve is only for demonstration purposes. A standard curve should be generated for each assay.

| Concentration (pg/ml) | Zero standard subtracted OD |
|-----------------------|-----------------------------|
| 0 | 0.000 |
| 31.25 | 0.034 |
| 62.5 | 0.055 |
| 125 | 0.087 |
| 250 | 0.192 |
| 500 | 0.421 |
| 1000 | 0.826 |
| 2000 | 1.568 |




Sensitivity

The minimum detectable dose of human IL20RA / IL20R1 was determined to be approximately **31.25 pg/ml**.

This is defined as at least three times standard deviations above the mean optical density of 10 replicates of the zero standard.

Precautions

The Stop Solution suggested for use with this Pair Set is an acid solution. Wear eye, hand, face, and clothing protection when using this material.



Trouble Shooting

| Problems | Possible Sources | Solutions |
|----------------------|---|---|
| No signal | Incorrect or no Detection Antibody was added | Add appropriate Detection Antibody and continue |
| | Substrate solution was not added | Add substrate solution and continue |
| | Incorrect storage condition | Check if the kit is stored at recommended condition and used before expiration date |
| Poor Standard Curve | Standard was incompletely reconstituted or was inappropriately stored | Aliquot reconstituted standard and store at -80 °C |
| | Imprecise / inaccurate pipetting | Check / calibrate pipettes |
| | Incubations done at inappropriate temperature, timing or agitation | Follow the general ELISA protocol |
| | Background wells were contaminated | Avoid cross contamination by using the sealer appropriately |
| Poor detection value | The concentration of antigen in samples was too low | Enriching samples to increase the concentration of antigen |
| | Samples were ineffective | Check if the samples are stored at cold environment. Detect samples in timely manner |
| High Background | Insufficient washes | Use multichannel pipettes without touching the reagents on the plate Increase cycles of washes and soaking time between washes |
| | TMB Substrate Solution was contaminated | TMB Substrate Solution should be clear and colorless prior to addition to wells |
| | Materials were contaminated. | Use clean plates, tubes and pipettes tips |
| Non-specificity | Samples were contaminated | Avoid cross contamination of samples |
| | The concentration of samples was too high | Try higher dilution rate of samples |