



User's Manual

Rib-P Ab ELISA Kit



DEIA4145



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: info@creative-diagnostics.com  Web: www.creative-diagnostics.com

PRODUCT INFORMATION

Intended Use

Anti-Rib-P is an indirect solid phase enzyme immunoassay (ELISA) for the quantitative measurement of IgG class autoantibodies against ribosomal P proteins in human serum or plasma. The assay is intended for research use only.

General Description

Antibodies to ribosomal proteins (anti-Rib-P or Anti-RPP) are of special clinical relevance in the differential diagnosis of systemic lupus erythematosus (SLE). Anti-Rib-P antibodies are directed against a common epitope of three phosphoproteins (P0, P1 and P2), which are major compounds of the 60s subunit of ribosomal RNP complexes. In cases of systemic lupus which are associated with depression and/or psychotic symptoms anti-Rib-P antibodies have been reported with an incidence of 40 to 90 percent. Determination of antibodies directed against ribosomal P-proteins is a valuable diagnostic tool for the differential diagnosis of SLE, especially in cases where classic SLE typical antibodies are not present.

Principles of Testing

Highly purified ribosomal P protein (Rib-P) is bound to microwells. Antibodies against this antigen, if present in diluted serum or plasma, bind to the respective antigen. Washing of the microwells removes unspecific serum and plasma components. Horseradish peroxidase (HRP) conjugated anti-human IgG immunologically detect the bound patient antibodies forming a conjugate/antibody/antigen complex. Washing of the microwells removes unbound conjugate. An enzyme substrate in the presence of bound conjugate hydrolyzes to form a blue color. The addition of an acid stops the reaction forming a yellow end-product. The intensity of this yellow color is measured photometrically at 450 nm. The amount of colour is directly proportional to the concentration of IgG antibodies present in the original sample.

Reagents And Materials Provided

1. Qty.1: Divisible microplate consisting of 12 modules of 8 wells each, coated with highly purified ribosomal P proteins P0, P1 and P2 (Rib-P). Ready to use.
2. 6 vials, 1.5 ml each: Anti-Rib-P Calibrators (A-F) in a serum/buffer matrix (PBS, BSA, NaN_3 <0.1% (w/w)) containing: IgG: 0; 12.5; 25; 50; 100; and 200 U/ml. Ready to use.
3. 2 vials, 1.5 ml each: Anti-Rib-P Controls in a serum/buffer matrix (PBS, BSA, NaN_3 <0.1% (w/w)) positive (1) and negative (2), for the respective concentrations see the enclosed QC insert. Ready to use.
4. 1 vial, 20 ml: Sample buffer (Tris, NaN_3 <0.1% (w/w)), yellow, concentrate (5 \times).
5. 1 vial, 15 ml: Enzyme conjugate solution (PBS, Proclin 300 <0.5% (v/v)), (light red) containing polyclonal rabbit anti-human IgG; labelled with horseradish peroxidase. Ready to use.
6. 1 vial, 15 ml: TMB substrate solution. Ready to use.
7. 1 vial, 15 ml Stop solution (contains acid). Ready to use.
8. 1 vial, 20 ml Wash solution (PBS, NaN_3 <0.1% (w/w)), concentrate (50 \times).

Materials Required But Not Supplied

1. Microplate reader capable of endpoint measurements at 450 nm
2. Multi-Channel Dispenser or repeatable pipet for 100 µl
3. Vortex mixer
4. Pipets for 10 µl, 100 µl and 1000 µl
5. Laboratory timing device
6. Data reduction software
7. Distilled or deionized water
8. Graduated cylinder for 100 and 1000 ml
9. Plastic container for storage of the wash solution

Storage

1. Store the kit at 2-8 °C.
2. Keep microplate wells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light during storage and usage.
5. Diluted sample buffer and wash buffer are stable for at least 30 days when stored at 2-8 °C.

Specimen Collection And Preparation

1. Collect whole blood specimens using acceptable medical techniques to avoid hemolysis.
2. Allow blood to clot and separate the serum by centrifugation.
3. Test serum should be clear and non-hemolyzed. Contamination by hemolysis or lipemia is best avoided, but does not interfere with this assay.
4. Specimens may be refrigerated at 2-8 °C for up to five days or stored at -20 °C up to six months.
5. Avoid repetitive freezing and thawing of serum samples. This may result in variable loss of autoantibody activity.
6. Testing of heat-inactivated sera is not recommended.

Reagent Preparation

Preparation of sample buffer

Dilute the contents of each vial of the sample buffer concentrate (5x) with distilled or deionized water to a final volume of 100 ml prior to use. Store refrigerated: stable at 2-8 °C for at least 30 days after preparation or until the expiration date printed on the label.

Preparation of wash solution

Dilute the contents of each vial of the buffered wash solution concentrate (50x) with distilled or deionized water to a final volume of 1000 ml prior to use. Store refrigerated: stable at 2-8 °C for at least 30 days after

preparation or until the expiration date printed on the label.

Sample preparation

Dilute all samples 1:100 with sample buffer before assay. Therefore combine 10 µl of sample with 990 µl of sample buffer in a polystyrene tube. Mix well.

Controls are ready to use and need not be diluted.

Assay Procedure

PROCEDURAL NOTES

1. Do not use kit components beyond their expiration dates.
2. Do not interchange kit components from different lots.
3. All materials must be at room temperature (20-28 °C).
4. Have all reagents and samples ready before start of the assay. Once started, the test must be performed without interruption to get the most reliable and consistent results.
5. Perform the assay steps only in the order indicated.
6. Always use fresh sample dilutions.
7. Pipette all reagents and samples into the bottom of the wells.
8. To avoid carryover contamination change the tip between samples and different kit controls.
9. It is important to wash microwells thoroughly and remove the last droplets of wash buffer to achieve best results.
10. All incubation steps must be accurately timed.
11. Control sera or pools should routinely be assayed as unknowns to check performance of the reagents and the assay.
12. Do not re-use microplate wells.

For all controls, the respective concentrations are provided on the labels of each vial. Using these concentrations a calibration curve may be calculated to read off the patient results semi-quantitatively.

TEST PROCEDURE

1. Prepare a sufficient number of microplate modules to accommodate controls and prediluted samples.
2. Pipet 100 µl of calibrators, controls and prediluted samples in duplicate into the wells.

	1	2	3	4	5	6
A	SA	SE	P1	P5		
B	SA	SE	P1	P5		
C	SB	SF	P2	P..		
D	SB	SF	P2	P..		
E	SC	C1	P3			
F	SC	C1	P3			
G	SD	C2	P4			
H	SD	C2	P4			

SA-SF: standards A to F
P1, P2....: patient sample 1, 2 ...
C1: positive control
C2: negative control

3. Incubate for 30 minutes at room temperature (20-28 °C).
4. Discard the contents of the microwells and wash 3 times with 300 µl of wash solution.

5. Dispense 100 µl of enzyme conjugate into each well.
6. Incubate for 15 minutes at room temperature.
7. Discard the contents of the microwells and wash 3 times with 300 µl of wash solution.
8. Dispense 100 µl of TMB substrate solution into each well.
9. Incubate for 15 minutes at room temperature.
10. Add 100 µl of stop solution to each well of the modules and incubate for 5 minutes at room temperature.
11. Read the optical density at 450 nm and calculate the results. Bi-chromatic measurement with a reference at 600-690 nm is recommended.

The developed colour is stable for at least 30 minutes. Read optical densities during this time.

Automation

The Anti-Rib-P ELISA is suitable for use on open automated ELISA processors. The test procedure detailed above is appropriate for use with or without automation.

Quality Control

This test is only valid if the optical density at 450 nm for Positive Control (1) and Negative Control (2) as well as for the Calibrator A and F complies with the respective range indicated on the Quality Control Certificate enclosed to each test kit ! If any of these criteria is not met, the results are invalid and the test should be repeated.

Calculation

For Anti-Rib-P a 4-Parameter-Fit with lin-log coordinates for optical density and concentration is the data reduction method of choice.

Recommended Lin-Log Plot: First calculate the averaged optical densities for each calibrator well. Use lin-log graph paper and plot the averaged optical density of each calibrator versus the concentration. Draw the best fitting curve approximating the path of all calibrator points. The calibrator points may also be connected with straight line segments. The concentration of unknowns may then be estimated from the calibration curve by interpolation.

Calculation example: The figures below show typical results for Anti-Rib-P ELISA. These data are intended for illustration only and should not be used to calculate results from another run.

No	Position	OD 1	OD 2	Mean	Conc. 1	Conc. 2	Mean	decl. Conc.	CV %
ST1	A 1/A 2	0.052	0.055	0.054	0.1	0.1	0.1	0.0	4
ST2	B 1/B 2	0.511	0.520	0.515	12.4	12.8	12.6	12.5	1
ST3	C 1/C 2	0.799	0.826	0.813	24	25	25	25	2
ST4	D 1/D 2	1.204	1.235	1.220	49	52	50	50	2
ST5	E 1/E 2	1.642	1.618	1.630	102	100	98	100	1
ST6	F 1/F 2	1.981	1.972	1.977	201	197	199	200	0

Interpretation Of Results

In a normal range study with serum samples from healthy blood donors the following ranges have been established with the Anti-Rib-P test:

negative: < 10 U/ml

positive: ≥ 10 U/ml

Precision

Statistics for coefficients of variation (CV) were calculated for each of three samples from the results of 24 determinations in a single run for Intra-Assay precision. Run-to-run precision was calculated from the results of 5 different runs with 6 determinations of each sample:

Intra-Assay		
Sample No.	Mean (IU/ml)	CV (%)
1	24.6	6.3
2	63.8	2.4
3	122.4	3.7

Inter-Assay		
Sample No.	Mean (IU/ml)	CV (%)
1	27.2	6.4
2	65.1	4.4
3	120.4	3.8

Sensitivity

The lower detection limit for Anti-Rib-P has been determined at 1.0 U/ml.

Specificity

The microplate is coated with highly purified ribosomal P protein (Rib-P) as antigen. The Anti-Rib-P test kits recognises only autoantibodies specific to the phospho-proteins P0, P1 and P2 of the 60s subunit of ribosomal RNP complexes. No cross reactivities to other RNP epitopes and autoantigens have been observed.

Interferences

No interference has been observed with haemolytic (up to 1000 mg/dL), lipemic (up to 3 g/dL triglycerides) or bilirubin (up to 40 mg/dL) containing sera. Nor have any interfering effects been observed with the use of anticoagulants. However for practical reasons it is recommended that grossly hemolyzed or lipemic samples should be avoided.

Precautions

1. All reagents of this kit are strictly intended for research use only.
2. Do not interchange kit components from different lots.
3. Components containing human serum were tested and found negative for HBsAg, HCV, HIV1 and HIV2 by FDA approved methods. No test can guarantee the absence of HBsAg, HCV, HIV1 or HIV2, and so all human serum based reagents in this kit must be handled as though capable of transmitting infection.
4. Avoid contact with the TMB (3,3',5,5'-Tetramethyl-benzidine). If TMB comes into contact with skin, wash thoroughly with water and soap.
5. Avoid contact with the Stop Solution which is acid. If it comes into contact with skin, wash thoroughly with water and seek medical attention.
6. Some kit components (i.e. Controls, Sample buffer and Buffered Wash Solution) contain Sodium Azide as

preservative. Sodium Azide (NaN_3) is highly toxic and reactive in pure form. At the product concentrations (0.09%), though not hazardous. Despite the classification as non-hazardous, we strongly recommend using prudent laboratory practices (see 8., 9., 10.).

7. Some kit components contain Proclin 300 as preservative. When disposing reagents containing Proclin 300, flush drains with copious amounts of water to dilute the components below active levels.
8. Wear disposable gloves while handling specimens or kit reagents and wash hands thoroughly afterwards.
9. Do not pipette by mouth.
10. Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled.
11. Avoid contact between the buffered Peroxide Solution and easily oxidized materials; extreme temperature may initiate spontaneous combustion.

Observe the guidelines for performing quality control in medical laboratories by assaying controls and/or pooled sera. During handling of all kit reagents, controls and serum samples observe the existing legal regulations.

