



## User's Manual

# MQCA ELISA Kit



DEIANS001



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

This kit can be used in quantitative and qualitative analysis of MQCA residue in Tissue (chicken, duck).

### General Description

Olaquinox is used to the feed additives of pork, chicken, and aquatic product, etc. Now the study shows that it has the carcinogenicity, teratogenicity, and Toxicity of photosensitive, so it has prohibition of use. The content of MQCA is the main mark of olaquinox residue. Analyze the MQCA is used to monitor the Olaquinox.

The kit is a new product based on ELISA technology, which is fast, easy, accurate and sensitive compared with common instrumental analysis and only needs 45 min in one detection, so it can considerably minimize operation error and work intensity.

### Principles of Testing

This kit is based on indirect-competitive ELISA technology. The microtiter wells are coated with coupling antigen. MQCA residue in the sample competes with the antigen coated on the microtiter plate for the antibody. After the addition of enzyme labeled anti-antibody, TMB substrate is used to show the color. Absorbance of the sample is negatively related to the MQCA residue in it, after comparing with the Standard Curve, multiplied by the dilution factor, MQCA residue quantity in the sample can be calculated.

### Reagents And Materials Provided

1. Microtiter plate with 96 wells coated with antigen
2. Standard solutions. (2 mLx6 bottles)  
0 ppb, 0.3 ppb, 0.9 ppb, 2.7 ppb, 8.1 ppb, 24.3 ppb
3. Spiking standard control: 1 mL, 100ppb
4. Concentrated Enzyme conjugate (1 mL), red cap
5. Antibody solution (7 mL), green cap
6. Solution A (7 mL), white cap
7. Solution B (7 mL), red cap
8. Stop solution (7 mL), yellow cap
9. 20xconcentrated wash solution (40 mL), transparent cap
10. 2xConcentrated extraction solution (50 mL), blue cap

### Materials Required But Not Supplied

1. Microtiter plate spectrophotometer (450nm/630 nm)
2. Homogenizer

3. Shaker
4. Centrifuge
5. Analytical balance (inductance: 0.01 g)
6. Graduated pipette: 10 mL
7. Rubber pipette bulb
8. Volumetric flask: 100 mL, 500 mL
9. Polystyrene centrifuge tube: 50 mL
10. Micropipettes: 20  $\mu$ L-200  $\mu$ L, 200  $\mu$ L-1000  $\mu$ L, 250  $\mu$ L-multipipette
11. N-hexane(AR)
12. Concentrated sulfuric acid (H<sub>2</sub>SO<sub>4</sub>, AR)
13. Ethyl acetate (AR)
14. Deionized water

## Storage

Storage condition: 2-8°C.

Storage period: 12 months.

## Specimen Collection And Preparation

### 1. Notice and precautions for before operation

- a. Please use one-off tips in the process of experiment, and change the tips when absorbing different reagent.
- b. Make sure that all experimental instruments are clean.
- c. Treated sample can't be stored.

### 2. Tissue sample(chicken, duck)

**Note: The whole operation should be as continuous as possible, after the sample is treated, please test as soon as possible.**

- a. Homogenize tissue samples with a homogenizer.
- b. Weigh  $2.0 \pm 0.05$  g homogenized sample to 50 ml Polystyrene centrifugal tube, add 2ml deionized water, shake for 2 min, add 8 mL ethyl acetate, 1 mL 2M H<sub>2</sub>SO<sub>4</sub> solution (solution 1). Shake for 1 min. then centrifuge at room temperature (20-25°C) for 5 min, at least 3000 g;
- c. Remove 4 mL supernatant organic phase to 10 mL clean dry glass tube, flow dry in nitrogen blow and water bath (50-60°C).
- d. Add 1 mL n-hexane, vortex for 30 s, add 1 mL extraction solution (solution 2), vortex for 2 min, mix completely, then centrifuge at room temperature (20-25°C) for 3 min, at least 3000 g;
- e. Remove the upper organic phase, take 100  $\mu$ L of the lower aqueous phase for assay.

### Dilution factor of samples:

Tissue: 1

## Reagent Preparation

### Solution 1: 2M H<sub>2</sub>SO<sub>4</sub> solution

Dissolve 10 mL 98.3% Concentrated sulfuric acid with deionized water and dilute to 100 mL;

### Solution 2: Extraction solution

Dilute 2xconcentrated extraction solution with deionized water in the volume ration of 1:1, which will be used for sample extraction. This solution can be stored for 1 month at 4°C.

### Solution 3: Wash solution

Dilute 20xConcentrated wash solution with deionized water in the volume ratio of 1:19, which will be used to wash the plates. This diluted solution can be stored for 1 month at 4°C.

## Assay Procedure

### 1. Notice before assay

- a. Make sure all reagents and microwells are all at room temperature (20-25°C).
- b. Return all the rest reagents to 2-8°C immediately after used.
- c. Washing the microwells correctly is an important step in the process of assay; it is the vital factor to the reproducibility of the ELISA analysis.
- d. Avoid the light and cover the microwells during incubation.

### 2. Assay Steps

- a. Take all reagents out at room temperature (20-25°C) for more than 30min, homogenize before use.
- b. Get the microwells needed out and return the rest into the zip-lock bag at 2-8°C immediately.
- c. The concentrated wash solution and concentrated extraction solution should be rewarmed to be at room temperature before use.
- d. **Number:** Numbered every microwell positions and all standards and samples should be run in duplicate. Record the standards and samples positions.
- e. **Add standard solution/sample:** Add 100 µL of standard solution or prepared sample to corresponding wells.
- f. **Mix the antibody solution and concentrated enzyme conjugate:** Mix the antibody solution and concentrated enzyme conjugate in the volume of 10:1.
- g. **Add mixed solution:** Add mixed solution 50 µL to each well, mix gently by rocking the plate manually and incubate for 30 min at 25°C with cover. Repeat the wash step again.
- h. **Wash:** Remove the cover gently and pour the liquid out of the wells and rinse the microwells with 250 µL diluted wash solution (**solution 3**) at interval of 10 s for 4-5 times. Absorb the residual water with absorbent paper (the rest air bubble can be eliminated with unused tip).
- i. **Coloration:** Add 50 µL solution A and 50 µL solution B to each well. Mix gently by rocking the plate manually and incubate for 15min at 25°C with cover(see **Precautions 8**).
- j. **Measure:** Add 50 µL the stop solution to each well. Mix gently by rocking the plate manually and measure

the absorbance at 450 nm (It's suggested measure with the dual-wavelength of 450/630 nm. Read the result within 5 min after addition of stop solution)

## Calculation

### 1. Percentage absorbance

The mean values of the absorbance values obtained for the standards and the samples are divided by the absorbance value of the first standard (zero standard) and multiplied by 100%. The zero standard is thus made equal to 100% and the absorbance values are quoted in percentages.

$$\text{Absorbance (\%)} = \frac{B}{B_0} \times 100\%$$

B —absorbance standard (or sample)

B<sub>0</sub> —absorbance zero standard

### 2. Standard Curve

- To draw a standard curve: Take the absorbance value of standards as y-axis, semi logarithmic of the concentration of the MQCA standards solution (ppb) as x-axis.
- The MQCA concentration of each sample (ppb), which can be read from the calibration curve, is multiplied by the corresponding Dilution factor of each sample followed, and the actual concentration of sample is obtained.

### 3. Please notice:

Special software has been developed for all data interpretation, which can be provided on request.

## Performance Characteristics

Accuracy: 100±20%

## Precision

Variation coefficient of the ELISA kit is less than 10%.

## Detection Limit

0.5 ppb

## Sensitivity

0.3 ppb

## Specificity

**Cross-reactions**

1. MQCA 100%
2. QCA <1%
3. Olaquinox <1%
4. Carbadox <1%

**Precautions**

1. The mean values of the absorbance values obtained for the standards and the samples will be reduced if the reagents and samples have not been regulated to room temperature (20-25°C).
2. Do not allow microwells to dry between steps to avoid unsuccessful reproducibility and operate the next step immediately after tap the microwells holder.
3. Shake each reagent gently before using.
4. Keep your skin away from the stop solution for it is the 2M H<sub>2</sub>SO<sub>4</sub> solution.
5. Don't use the kits out of date. Don't exchange the reagents of different batches, or else it will drop the sensitivity.
6. Keep the ELISA kits at 2-8°C, do not freeze. Seal rest microwell plates, avoid sunlight during all incubations. Covering the microtiter plates is recommended.
7. Substrate solution should be abandoned if it turns colors. The reagents may be turn bad if the absorbance value (450/630 nm) of the zero standard is less than 0.5 (A<sub>450 nm</sub><0.5).
8. The coloration reaction need 10-15min after the addition of solution A and solution B; But you can prolong the incubation time ranges from 20 min to more if the color is too light to be determined, never exceed 25 min, on the contrary, shorten the incubation time properly.
9. The optimal reaction temperature is 25°C. Higher or lower temperature will lead to the changes of sensitivity and absorbance values.

