



## User's Manual

# T7 RNA Polymerase ELISA Kit

REF

DEIANS032



96T

RUO

This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

---

### Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)  Web: [www.creative-diagnostics.com](http://www.creative-diagnostics.com)

---

## PRODUCT INFORMATION

### Intended Use

The kit is used for detection and quantitative determination of T7 RNA Polymerase in the samples.

### Principles of Testing

This kit uses the double antibody sandwich method: The microwell plate has been pre-coated with anti-T7 RNA polymerase antibody. The sample is added and incubated with horseradish peroxidase (HRP) labeled antibody to form antibody-antigen-enzyme-labeled antibody complex. After washing, Then TMB substrate solution is added and catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The depth of the color is positively correlated with the T7 RNA polymerase content in the sample. The absorbance (OD value) is measured at 450nm.

### Reagents And Materials Provided

1. ELISA microplate, 96T
2. Standard concentrate (0.1mg/ml), 0.05mL
3. Enzyme conjugate (100×), 0.08mL
4. Sample dilution buffer, 30mL
5. Enzyme conjugate dilution buffer, 8mL
6. Wash concentrate (10×), 30mL
7. Substrate solution, 12mL
8. Stop solution, 6mL
9. Sealing film, 4 films
10. Manual, 1 copy

### Materials Required But Not Supplied

1. Microplate reader with 450±10nm filter(better if can detect at 450 and 650 nm wavelength).

### Storage

1. Unopened kit should be stored at 2~8°C for stability in shelf life.

### Reagent Preparation

1. Bring all reagents to room temperature (23°C±2°C) before use.
2. Dilute concentrated washing liquid 10× with purified water according to 1:9 volume ratio into washing working liquid.

- Dilute Enzyme conjugate (100×) to the working concentration by enzyme conjugate dilution buffer before use.
- Dilute the Standard concentrate (0.1mg/ml) with Sample dilution buffer to 32ng/mL, 16ng/mL, 8ng/mL, 4ng/mL, 2ng/mL, 1ng/mL, 0.5ng/mL, 0ng/mL. We recommend standards can be diluted as following chart:

ID	Concn. (ng/mL)	Dilution buffer	Antigen
A0	1000	0.198mL	2 $\mu$ L Standard concentrate
A	32	0.242mL	8 $\mu$ L A0
B	16	0.125mL	0.125mL of A
C	8	0.125mL	0.125mL of B
D	4	0.125mL	0.125mL of C
E	2	0.125mL	0.125mL of D
F	1	0.125mL	0.125mL of E
G	0.5	0.125mL	0.125mL of F
H	0	0.125mL	/

- Mix 1X enzyme conjugate solution with standard product and sample in 1:1 vortex and then use.  
Recommended dilution method: directly add an equal amount of 1× enzyme conjugate solution to the above prepared standard solution; Or take a new centrifuge tube and mix the standard solution and 1× enzyme conjugate solution 1:1.

## Assay Procedure

- Determine the number of strips required for the assay. Insert the strips in the frames for use. Remaining plate strips not used in this assay should be repacked in the bag with desiccant. Close the bag tightly for refrigerated storage.
- It is recommended that standard products and samples should be tested with multiple wells. Add the standard product/sample and 1X enzyme conjugate mixture into the microhole of the enzyme label plate according to 100 $\mu$ L/ well.
- Cover with the plate sealer. and incubate at 37°C for 60min.
- Discard the liquid and pat it dry on a clean towel or absorbent paper. Fill each well with washing liquid (300 $\mu$ L) and let it stand for 15-30s. Shake off the washing liquid and pat it dry. Repeat washing the board three times.
- Add 100 $\mu$ L of TMB substrate solution to each well. Cover with the plate sealer. Incubate at 37°C for 15 min away from light.
- Add 50 $\mu$ L of the Stop solution into each well. Read absorbance at 450/650nm.



## Calculation

1. Calculate the average light absorption value of the standard, blank control, and sample. The standard curve is drawn with the concentration of the standard product as the horizontal(X) coordinate and the average light absorption value of the standard product as the vertical(Y) coordinate. (The light absorption value in dualwavelength detection mode is 450nm minus 650nm).
2. It is recommended to perform the calculation with computer-based curve-fitting software such as curve expert 1.3 or ELISA Calc in a 4 parameter non-linear fit model.

## Precision

Intra-assay CV  $\leq 10\%$ , Inter-assay CV  $\leq 15\%$

## Detection Limit

1. Lower limit of detection:  $< 0.1$  ng/mL
2. Lower limit of quantitation: 0.25 ng/mL

## Specificity

Only specific recognition of T7 RNA polymerase, other IVT tool enzymes do not interfere with the measurements.

## Linearity

0.25~32 ng/mL

## Recovery

80%~120%

## Precautions

1. TMB reaction temperature and time are critical, please control them according to the instruction strictly.
2. During the washing process, the washing liquid should be soaked in the plate for 15~30s and then dried to fully wash the non-specific adsorbed components..
3. All the reagents should be mixed thoroughly prior to use and avoid bubbles during sample or reagents addition.
4. If crystals have formed in the concentrated wash buffer(10 $\times$ ), warm to 37°C and mix gently until the crystals are completely dissolved.
5. Avoid assay of samples containing sodium azide (NaN<sub>3</sub>), which will destroy the horseradish peroxidase activity and make the detection value low.



6. Different batches of reagents cannot be mixed. In addition, the pipette tips should not be mixed to avoid cross contamination.
7. For your safety and health, please wear a lab coat and wear gloves for experimental operation.
8. This product is intended for research use only, and shall not be used for clinical medical diagnosis and other irrational purposes.

