



User's Manual

Endotoxin ELISA Kit



DEIAPY07261



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: info@creative-diagnostics.com  Web: www.creative-diagnostics.com

PRODUCT INFORMATION

General Description

Endotoxins are a component of the cell wall of certain bacteria. Bacteria which contain endotoxin are known as gram negative bacteria. The endotoxins are released from the bacteria as they are destroyed and elicit a strong response from the immune system by activating various components of the immune system and inducing cytokine release.

Principles of Testing

The Endotoxin ELISA Kit is based on a competitive enzyme immunoassay technique. The microtiter well-plate in this kit has been pre-coated with an anti- Endotoxin antibody. Sample or standards are added to the wells along with a fixed quantity of biotinylated Endotoxin and incubated. The Endotoxin found in the sample or standards competes with the biotinylated Endotoxin for limited binding sites on the immobilized anti-Endotoxin antibody. Excess unbound biotinylated Endotoxin and sample or standard Endotoxin is washed from the plate. Avidin-HRP conjugate is added, incubated and washed. An enzymatic reaction is then produced through the addition of TMB substrate which is catalyzed by the immobilized HRP to generate a blue color product that changes to yellow after adding acidic stop solution. The density of yellow coloration is measured by reading the absorbance at 450 nm which is quantitatively proportional to the amount of biotinylated Endotoxin captured in the well and inversely proportional to the amount of Endotoxin which was contained in the sample or standard.

Reagents And Materials Provided

1. Anti-Endotoxin Microplate 96 Wells (12 × 8 Well strips)
2. Endotoxin Lyophilized Standard 2 × 1 EU
3. Sample Diluent 1 × 20 mL
4. 100× Endotoxin-Biotin Complex 1 × 60 µL
5. 100× Avidin-HRP Conjugate 1 × 120 µL
6. Biotin Complex Diluent 1 × 12 mL
7. Conjugate Diluent 1 × 12 mL
8. 25× Wash Buffer 1 × 30 mL
9. TMB Substrate 1 × 10 mL
10. Stop Solution 1 × 10 mL

Materials Required But Not Supplied

1. Microplate reader capable of reading absorbance at 450 nm.
2. Automated plate washer (optional).
3. Pipettes capable of precisely dispensing 0.5 µL through 1 mL volumes of aqueous solutions.
4. Pipettes or volumetric glassware capable of precisely measuring 1 mL through 100 mL of aqueous

solutions.

5. New, clean tubes and/or micro-centrifuge tubes for the preparation of standards or samples.
6. Absorbent paper or paper toweling.
7. Distilled or deionized ultrapure water.
8. 37°C Incubator (optional)

Storage

Open kit immediately upon receipt. Store components at -20°C (NOTE: exceptions below) for 6 months or until expiration date. Avoid any freeze/thaw cycles.

Specimen Collection And Preparation

- Samples must be tested to determine if the kit is valid.
- Store samples to be assayed at 4°C for 24 hours prior being assayed.
- For long term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

General Sample Preparation Guidelines

1. **Serum:** Use a serum separator tube (SST) and allow samples to clot for 30 minutes at room temperature or overnight at 4°C before centrifugation for 15 minutes at 1,000 x g. Remove serum and assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.
2. **Plasma:** Collect plasma using EDTA, or heparin as an anticoagulant. Centrifuge for 15 minutes at 1,000 x g at 4°C within 30 minutes of collection. Assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze-thaw cycles.
3. **Tissue Homogenates:** Rinse 100 mg of tissue with 1x PBS, then homogenize in 1 mL of 1x PBS and store overnight at -20°C. Perform two freeze-thaw cycles to break the cell membranes, then centrifuge the homogenate for 5 minutes at 5,000 xg, 2-8°C. Remove the supernatant and assay immediately. Alternatively, aliquot and store samples at -20°C or -80°C. Centrifuge the sample again after thawing before the assay. Avoid repeated freeze-thaw cycles.
4. **Cell Lysates:** Adherent cells should be detached with trypsin and then collected by centrifugation (suspension cells can be collected by centrifugation directly). Wash cells three times in cold PBS. Resuspend cells in PBS (1x) and ultrasonicate the cells 4 times (or freeze cells at $\leq -20^{\circ}\text{C}$. Thaw cells with gentle mixing. Repeat the freeze/thaw cycle 3 times.) Centrifuge at 1,500 xg for 10 minutes at 2 - 8°C to remove cellular debris.
5. **Cell culture supernatants and other biological fluids:** Remove particulates by centrifugation and assay immediately or aliquot and store samples at -20°C or -80°C. Avoid repeated freeze/thaw cycles.

Recombinant Proteins: Due to the possibility of mismatching between antigen from other origin and antibody used in our kits (e.g. antibody targets conformational epitope rather than linear epitope), some native or recombinant proteins from other manufacturers may not be recognized by our products

Sample Dilution

Target protein concentration must be estimated and appropriate sample dilution selected such that the final target protein concentration falls near the middle of the assay linear dynamic range. Samples exhibiting

saturation should be further diluted.

1. Prior to performing the full experiment, test a serially diluted representative sample.
 - Alternately, a small pool of several samples can also be used with this same method if sample volume is limited.

-or-

- Refer to published literature for expected concentrations and derive the optimal dilution level based on the expected dynamic range of the kit.
2. Dilute samples using Sample Diluent.
 3. Mix diluted samples gently and thoroughly.
 4. Pipetting less than 2 μ L is not recommended for optimal assay accuracy.

Reagent Preparation

Technical Application Tips

1. Do not mix or substitute components from other kits.
2. To ensure the validity of experimental operation, it is recommended that pilot experiments using standards and a small selection of sample dilutions to ensure optimal dilution range for quantitation.
3. Samples exhibiting OD measurements higher than the highest standard should be diluted further in the appropriate sample dilution buffers.
4. Prior to using the kit, briefly spin component tubes to collect all reagents at the bottom.
5. Replicate wells are recommended for standards and samples.
6. Cover microplate while incubating to prevent evaporation.
7. Do not allow the microplate wells dry at any point during the assay procedure.
8. Do not reuse tips or tube to prevent cross contamination.
9. Avoid causing bubbles or foaming when pipetting, mixing or reconstituting.
10. Completely remove of all liquids when washing to prevent cross contamination.
11. Prepare reagents immediately prior to use and do not store, with the exception of the top standard.
12. Equilibrate all materials to ambient room temperature prior to use (standards exception).
13. For optimal results for inter- and intra-assay consistency, equilibrate all materials to room temperature prior to performing assay (standards exception) and perform all incubations at 37°C.
14. Pipetting less than 1 μ L is not recommended for optimal assay accuracy.
15. Once the procedure has been started, all steps should be completed without interruption. Ensure that all reagents, materials and devices are ready at the appropriate time.
16. Incubation times will affect results. All wells should be handled in the same sequential order and time intervals for optimal results.
17. Samples containing precipitates or fibrin strands or which are hemolytic or lipemic might cause inaccurate results due to interfering factors.
18. TMB Substrate is easily contaminated and should be colorless or light blue until added to plate. Handle carefully and protect from light.

Preparation

Equilibrate all materials to room temperature prior to use and use immediately.

Standards

1. Prepare a fresh standard curve for each assay performed. Reconstituted standards cannot be stored for later use.

1. 0 EU/ml: Reconstitute 1 tube of lyophilized Standard with 1.0 ml of Sample Diluent. Incubate at room temperature for 10 minutes with gentle agitation (avoid foaming).

D1 (1 EU/ml): Pipette 500 µL of Stock Standard into 0 µL of Sample Diluent

D2 (0.5 EU/ml): Pipette 250 µL of D1 into 250 µL of Sample Diluent

D3 (0.25 EU/ml): Pipette 250 µL of D2 into 250 µL of Sample Diluent

D4 (0.125 EU/ml): Pipette 250 µL of D3 into 250 µL of Sample Diluent

D5 (0.063 EU/ml): Pipette 250 µL of D4 into 250 µL of Sample Diluent

D6 (0.032 EU/ml): Pipette 250 µL of D5 into 250 µL of Sample Diluent

D7 (0.016 EU/ml): Pipette 250 µL of D6 into 250 µL of Sample Diluent

Zero Standard (0 EU/ml): Use Sample Diluent alone

1x Endotoxin-Biotin Complex

1. Prepare the 1X Endotoxin-Biotin Complex immediately prior to use by diluting the 100X Endotoxin-Biotin Complex 1:100 with Complex Diluent.
2. For each well strip to be used in the experiment (8-wells) prepare 500 µL by adding 5 µL of 100x Endotoxin-Biotin Complex to 495 µL Complex Diluent.
3. Mix thoroughly and gently. Hold no longer than 2 hours prior to using in procedure. Do not store at 1x concentration for future use.

1x Avidin-HRP Conjugate

1. Prepare the 1x Avidin-HRP Conjugate immediately prior to use by diluting the 100x Avidin-HRP Conjugate 1:100 with Conjugate Diluent as follows.
2. Briefly and gently mix the 100x Avidin-HRP Conjugate prior to pipetting.
3. For each well strip to be used in the experiment (8-wells) prepare 1,000 µL 1x Avidin-HRP Conjugate by adding 10 µL of 100x Avidin-HRP Conjugate to 990 µL Conjugate Diluent.
4. Mix thoroughly and gently. Hold no longer than 2 hours prior to using in procedure.

Microplate Preparation

1. Micro-plates are provided ready to use and do not require rinsing or blocking.
2. Unused well strips should be returned to the original packaging, sealed and stored at 4°C.
3. Equilibrate microplates to ambient temperatures prior to opening to reduce potential condensation.

1x Wash Buffer

1. If crystals have formed in the 25x Wash Buffer concentrate, equilibrate to room temperature and mix gently until crystals have completely dissolved.
2. Add the entire 15 mL contents of the 25x Wash Buffer bottle to 360 mL of ultra-pure water to a clean >

1,000 mL bottle or other vessel.

3. Seal and mix gently by inversion. Avoid foaming or bubbles.
4. Store the 1× Wash Buffer at room temperature until ready to use in the procedure. Store the prepared 1× Wash Buffer at 4°C for no longer than 1 week. Do not freeze.

Assay Procedure

- Equilibrate all reagents and materials to ambient room temperature prior to use in the procedure.
 - Optimal results for intra- and inter-assay reproducibility will be obtained when performing incubation steps at 37°C as indicated below.
1. Determine the required number of wells and return any remaining unused wells and desiccant to the pouch.
 2. Retain at least one well as an absolute Blank without any samples or reagents.
 3. Add 50 µL of serially titrated standards, diluted samples or blank into wells of the Endotoxin Microplate. At least two replicates of each standard, sample or blank is recommended.
 4. Immediately add 50 µL of 1× Endotoxin-Biotin Complex to each well (excluding absolute Blank).
 5. Cover the plate with the plate sealer and incubate for 60 minutes.
 6. Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
 7. Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
 8. Wash plate three times with 1× Wash Buffer as follows:
 - 8.1 Add 200 µL of 1× Wash Buffer to each assay well.
 - 8.2 Incubate for 2 minutes.
 - 8.3 Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
 - 8.4 Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
 - 8.5 Repeat steps 8.1 through 8.4 two more times.
 9. Add 100 µL of 1× Avidin-HRP Conjugate to each well.
 10. Cover the plate with the plate sealer and incubate for 45 minutes.
 11. Discard the liquid in the wells by rigorously flicking into an acceptable waste receptacle or aspiration.
 12. Gently blot any remaining liquid from the wells by tapping inverted on the benchtop onto paper toweling. Do not allow the wells to completely dry at any time.
 13. Repeat wash as in step 8.
 14. Add 90 µL of TMB Substrate to each well, cover with plate sealer and incubate at 37°C in the dark for 15-30 minutes. Wells should change to gradations of blue. If the color is too deep based on the standard, adjust incubation times. (NOTE: optimal incubation time must be determined by the user. Optimal development can be visualized by blue shading in the bottom four standard wells, while the remaining standards still appear clearer.)
 15. Add 50 µL of Stop Solution to each well. Well color should change to gradations of yellow immediately. Add



the Stop Solution in the same well order as done for the TMB Substrate.

16. Read the O.D. absorbance at 450 nm with a standard microplate reader within 5 minutes of stopping the reaction in step 10.15. If wavelength correction is available, set to 540 nm or 570 nm.

Calculation

For analysis of the assay results, calculate the Relative OD₄₅₀ for each test or standard well as follows:

$$(\text{Relative OD}_{450}) = (\text{Well OD}_{450}) - (\text{Mean Blank Well OD}_{450})$$

The standard curve is generated by plotting the mean replicate Relative OD₄₅₀ of each standard serial dilution point vs. the respective standard concentration. The concentration contained in the samples can be interpolated by using linear regression of each mean sample Relative OD₄₅₀ against the standard curve. This is best achieved using curve fitting software. A standard curve should be generated each time the test is performed.

Note: if wavelength correction readings were available, subtract the readings at 540 nm or 570 nm from the readings at 450 nm. This may provide greater reading accuracy.

Note: if the samples measured were diluted, multiply the derived mean sample concentration by the dilution factor for a final sample concentration.

Performance Characteristics

Intra-assay CV%: < 4.3% (n = 20)

Inter-assay CV%: < 5.5% (n = 20)

Sensitivity

0.005 EU/mL

Specificity

No detectable cross-reactivity with other relevant proteins

Precautions

1. Read instructions fully prior to beginning use of the assay kit.
2. Any deviations or modifications from the described method or use of other reagents could result in a reduction of performance.
3. Reduce exposure to potentially harmful substances by wearing personal protective lab equipment including lab coats, gloves and glasses.
4. Kit cannot be used beyond the expiration date on the label.