



User's Manual

Durvalumab ELISA Kit



DEIAZ0027



96T



This product is for research use only and is not intended for diagnostic use.

For illustrative purposes only. To perform the assay the instructions for use provided with the kit have to be used.

Creative Diagnostics

 Address: 45-1 Ramsey Road, Shirley, NY 11967, USA

 Tel: 1-631-624-4882 (USA) 44-161-818-6441 (Europe)  Fax: 1-631-938-8221

 Email: info@creative-diagnostics.com  Web: www.creative-diagnostics.com

PRODUCT INFORMATION

Intended Use

Enzyme immunoassay for quantitative analysis of free denosumab in serum and plasma samples

General Description

Denosumab is a novel, fully human IgG2 monoclonal antibody specific to receptor activator of nuclear factor kappa-B ligand (RANKL), suppresses bone resorption markers in patients with a variety of metastatic tumours and is being investigated in multiple clinical trials for the prevention and treatment of bone metastases. Chemically, it consists of 2 heavy and 2 light chains. Each light chain consists of 215 amino acids. Each heavy chain consists of 448 amino acids with 4 intramolecular disulphides. Denosumab is designed to target RANKL (RANK ligand), a protein that acts as the primary signal to promote bone removal/resorption. In many bone losses, RANKL overwhelms the body's natural defence against bone destruction. Denosumab prevents RANKL from activating its receptor, RANK, on the surface of osteoclasts and their precursors. Prevention of the RANKL/RANK interaction inhibits osteoclast formation, function, and survival, thereby decreasing bone resorption and increasing bone mass and strength in both cortical and trabecular bone.

Principles of Testing

Solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle. Standards and samples (serum or plasma) are incubated in the microtiter plate coated with the reactant for denosumab. After incubation the wells are washed. Then, horse radish peroxidase (HRP) conjugated probe is added and binds to denosumab captured by the reactant on the surface of the wells. Following incubation wells are washed and the bound enzymatic activity is detected by addition of tetramethylbenzidine (TMB) chromogen substrate. Finally, the reaction is terminated with an acidic stop solution. The color developed is proportional to the amount of denosumab in the sample or standard. Results of samples can be determined directly using the standard curve.

Reagents And Materials Provided

1. Microtiter Plate: 1 × 12 × 8, Break apart strips. Microtiter plate with 12 rows each of 8 wells coated with reactant.
2. Standard A-E: 1 mL (each)
Standard A: 300 ng/mL
Standard B: 100 ng/mL
Standard C: 30 ng/mL
Standard D: 10 ng/mL
Standard E: 0 ng/mL

Ready to use. Used for the standard curve and control. Contains Durvalumab, human serum and

stabilizer, < 0,1 % NaN₃.

3. Control low and high levels: 1 mL (each) Ready to use. Contains human serum and stabilizer, < 0,1 % NaN₃. Control concentrations are given in "Quality control certificate".
4. Assay Buffer: 1 × 50 mL, Ready to use. Blue colored. Contains proteins, < 0,1 % NaN₃.
5. Conjugate: 1 × 12 mL, Horse radish peroxidase. Conjugated probe ready to use. Red colored. Contains HRP conjugated probe, stabilizer and preservatives.
6. Substrate: 1 × 12 mL, TMB substrate solution. Ready to use. Contains 3,3',5,5' - Tetramethylbenzidine (TMB).
7. Stop Buffer: 1 × 12 mL, Ready to use. 1N HCl.
8. Wash Buffer (20×): 1 × 50 mL, Prepared concentrated (20×) and should be diluted with the dilution rate given in the "Reagent Preparation" before the test. Contains buffer with tween 20.
9. Foil: 2 × 1, For covering microtiter plate during incubation.

Materials Required But Not Supplied

1. Micropipettes and tips
2. Calibrated measures
3. Tubes for sample dilution
4. Wash bottle, automated or semi-automated microtiter plate washing system
5. Microtiter plate reader capable of measuring optical density with a photometer at OD 450 nm with reference wavelength 650 nm (450/650 nm)
6. Distilled or deionised water, paper towels, pipette tips and timer

Storage

The kit is shipped at ambient temperature (10-30°C) and should be stored at 2-8°C for long term storage. Keep away from heat or direct sunlight. The strips of microtiter plate are stable up to the expiry date of the kit in the broken, but tightly closed bag when stored at 2-8°C.

Specimen Collection And Preparation

The usual precautions for venipuncture should be observed. Do not use grossly haemolytic, icteric or lipemic specimens. Samples appearing turbid should be centrifuged before testing to remove any particulate material. Avoid repeated freezethaw cycles for serum/plasma samples.

Samples should be diluted with the dilution rate given in the "Reagent Preparation" before the test.

Drug infusions may camouflages/mask the presence of antibody to drugs in serum/plasma samples. Therefore, blood sampling time is critical for detection of antibodies. It is recommended to take the blood sample just before the scheduled dose (trough specimen).

Stability (serum/plasma):

2-8°C, 2 days

-20°C, 6 months

Reagent Preparation

Procedure Notes:

1. Any improper handling of samples or modification of the test procedure may influence the results. The indicated pipetting volumes, incubation times, temperatures and pre-treatment steps must be performed strictly according to the instructions. Use calibrated pipettes and devices only.
2. Once the test has been started, all steps should be completed without interruption. Make sure that required reagents, materials and devices are prepared ready at the appropriate time. Allow all reagents and specimens to reach room temperature (18-25°C) and gently swirl each vial of liquid reagent and sample before use. Mix reagents without foaming.
3. Avoid contamination of reagents, pipettes and wells/tubes. Use new disposable plastic pipette tips for each reagent, standard or specimen. Do not interchange caps. Always cap not used vials. Do not reuse wells/tubes or reagents.
4. Use a pipetting scheme to verify an appropriate plate layout.
5. Incubation time affects results. All wells should be handled in the same order and time sequences. It is recommended to use an eight-channel micropipette for pipetting of solutions in all wells.
6. Microplate washing is important. Improperly washed wells will give erroneous results. It is recommended to use a multichannel pipette or an automatic microplate washing system. Do not allow the wells to dry between incubations. Do not scratch coated wells during rinsing and aspiration. Rinse and fill all reagents with care. While rinsing, check that all wells are filled precisely with wash buffer, and that there are no residues in the wells.
7. Humidity affects the coated wells/tubes. Do not open the pouch until it reaches room temperature. Unused wells/tubes should be returned immediately to the resealed pouch including the desiccant.

Component:

Wash buffer (Must be prepared before starting assay procedure)

Dilute: 10 mL (e.g.)

With Distilled water: Up to 200 mL

Dilution Ratio: 1/20

Remarks: Warm up 37°C to dissolve crystals. Mix vigorously

Storage: 2-8°C

Stability: 2 weeks

Dilution of samples

Sample: Serum/Plasma

Diluent: Assay buffer

Dilution Ratio: 1/1000

Remarks: For final dilution 1/1000

- First step: 1/10 dilution

10 µL sample + 90 µL assay buffer

– Second step: 1/100 dilution

5 µL sample (1/10 diluted) + 495 µL assay buffer

Patient samples with a concentration of drug above the measuring range are to be rated as > "Highest Standard (Standard A)". The result must not be extrapolated. The patient sample in question should be further diluted with assay buffer and retested.

Assay Procedure

Total assay time: 70 minutes

1. Pipette 100 µL of each "Standards", "Low level control", "High level control" and diluted samples into the respective wells of microtiter plate

Wells

A1: Standard A

B1: Standard B

C1: Standard C

D1: Standard D

E1: Standard E

F1: Low level control

G1: High level control

2. - Cover the plate with adhesive foil
- Briefly mix contents by gently shaking the plate
- Incubate 30 minutes at room temperature (18-25°C)
3. - Remove adhesive foil
- Discard incubation solution
- Wash plate three times each with 300 µL "Wash Buffer"
- Remove excess solution by tapping the inverted plate on a paper towel
4. Pipette 100 µL "Conjugate" into each well
5. - Cover the plate with adhesive foil
- Incubate 30 minutes at room temperature (18-25°C)
6. - Remove adhesive foil
- Discard incubation solution
- Wash plate three times each with 300 µL "Wash Buffer"
- Remove excess solution by tapping the inverted plate on paper towel
7. - Pipette 100 µL "Substrate" into each well
8. - Incubate 10 minutes without adhesive foil at room temperature (18-25°C) in the dark
9. - Stop the substrate reaction by adding 100 µL "Stop Solution" into each well. Briefly mix contents by gently



shaking the plate color changes from blue to yellow

10. Measure optical density with a photometer at OD 450 nm with reference wavelength 650 nm (450/650 nm) within 30 minutes after pipetting the "Stop Solution"

The ELISA is also suitable to run on automated ELISA processors.

Quality Control

The test results are only valid if the test has been performed following the instructions. Moreover, the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable standards/laws. For a valid study, the OD 450/650 of the highest standard should be > 1.500 and the OD 450/650 of the lowest standard should be < 0.150 . In case of any deviation the following technical issues (but not limited to) should be reviewed:

Expiration dates of reagents, storage conditions, pipettes, devices, incubation conditions, washing methods, etc.

Calculation

1. Create a standard curve by using the standards. OD 450/650 nm for each standard on the vertical (Y-axis) axis versus the corresponding drug concentration on the horizontal (X-axis) axis.
2. The concentration of the samples can be read directly from this standard curve. Using the absorbance value for each sample, determine the corresponding concentration of drug from the standard curve. Find the absorbance value on the Y-axis and extend a horizontal line to the curve. At the point of intersection, extend a vertical line to the X-axis and read the drug concentration of the unknown sample.
3. If computer data is going to be used, we recommend primarily "Four Parameter Logistic (4PL)" or secondly the "point-to-point calculation".
4. To obtain the exact values of the samples, the concentration determined from the standard-curve must be multiplied by the dilution factor ($1000\times$). Any sample reading greater than the highest standard should be further diluted appropriately with assay buffer and retested. Therefore, if the pre-diluted samples have been further diluted, the concentration determined from the standard curve must be multiplied by the further dilution factor.

e.g.; If the pre-diluted sample further diluted in a ratio of 1/10 then results should be multiplied by 10000.

5. For low and high level controls values, refer to "Quality Control Certificate" provided by each kit.

Precision

Intra-assay and inter-assay CVs $< 30\%$

Sensitivity

The lowest detectable level (Lowest detection limit, LOD) that can be distinguished from the zero standard is 6 ng/mL

Functional sensitivity (Limit of quantification-LOQ): 10 ng/mL

Specificity

There is no cross reaction with native serum immunoglobulin

Recovery

< 100 ± 30%

Precautions

1. For professional use only.
2. Obey lot number and expiry date. Do not mix reagents of different lots. Do not use expired reagents.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood. For further information (clinical background, test performance, automation protocols, alternative applications, literature, etc.) please refer to the local distributor.
4. Follow good laboratory practice and safety guidelines. Wear lab coats, disposable latex gloves and protective glasses where necessary.
5. All reagents of this kit containing human serum or plasma (standards etc.) have been tested and were found negative for HIV I/II, HBsAg and Anti-HCV. However, a presence of these or other infectious agents cannot be excluded absolutely and therefore reagents should be treated as potential biohazards in use and for disposal.
6. Reagents of this kit containing hazardous material may cause eye and skin irritations.
7. Chemicals and prepared or used reagents must be treated as hazardous waste according the national biohazard safety guidelines or regulations.